

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111022\
 Data File : VU051804.D
 Acq On : 10 Nov 2022 16:34
 Operator : JC/MD
 Sample : N5571-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBMW5

Quant Time: Nov 11 00:40:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 11 00:38:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	208355	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	212074	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	81398	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	69117	4.405	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.000%
7) Chloroethane-d5	1.906	69	66888	4.087	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.800%
11) 1,1-Dichloroethene-d2	2.562	65	28730	5.045	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.800%
20) 2-Butanone-d5	4.616	46	202911	35.704	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	71.400%
24) Chloroform-d	5.060	84	145748	3.969	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.400%
26) 1,2-Dichloroethane-d4	5.697	65	82016	4.135	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.600%
32) Benzene-d6	5.726	84	304066	4.543	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.800%
36) 1,2-Dichloropropane-d6	6.690	67	101845	4.352	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.000%
41) Toluene-d8	7.896	98	240557	4.571	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.400%
43) trans-1,3-Dichloroprop...	8.179	79	31474	4.399	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.000%
46) 2-Hexanone-d5	8.629	63	160778	37.326	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	74.660%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	92698	4.323	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.400%
66) 1,2-Dichlorobenzene-d4	12.192	152	87842	5.880	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111022\
Data File : VU051804.D
Acq On : 10 Nov 2022 16:34
Operator : JC/MD
Sample : N5571-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBMW5

Quant Time: Nov 11 00:40:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 11 00:38:34 2022
Response via : Initial Calibration

