

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111023\
 Data File : VU056166.D
 Acq On : 10 Nov 2023 12:52
 Operator : MD/SY
 Sample : 05297-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Nov 10 22:20:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 10 22:17:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	396591	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	397841	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	180017	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	91915	3.734	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.600%
7) Chloroethane-d5	1.911	69	69102	4.136	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.800%
11) 1,1-Dichloroethene-d2	2.563	65	41420	3.819	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.400%
20) 2-Butanone-d5	4.624	46	181635	44.976	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.960%
24) Chloroform-d	5.055	84	185750	3.964	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.200%
26) 1,2-Dichloroethane-d4	5.698	65	85119	4.125	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.600%
32) Benzene-d6	5.721	84	378224	3.958	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.200%
36) 1,2-Dichloropropane-d6	6.685	67	112688	4.006	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.200%
41) Toluene-d8	7.894	98	333372	3.807	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.200%
43) trans-1,3-Dichloroprop...	8.177	79	39787	3.701	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.000%
46) 2-Hexanone-d5	8.631	63	154354	43.802	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.600%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	82648	4.117	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.400%
66) 1,2-Dichlorobenzene-d4	12.190	152	117417	4.230	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	84.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111023\
Data File : VU056166.D
Acq On : 10 Nov 2023 12:52
Operator : MD/SY
Sample : 05297-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Nov 10 22:20:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 10 22:17:51 2023
Response via : Initial Calibration

