

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111023\
 Data File : VU056192.D
 Acq On : 11 Nov 2023 00:13
 Operator : MD/SY
 Sample : 05339-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCDJ2

Quant Time: Nov 11 01:44:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 11 00:40:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	363743	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	366174	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	160634	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	98280	4.353	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	87.000%	
7) Chloroethane-d5	1.908	69	79617	5.196	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	104.000%	
11) 1,1-Dichloroethene-d2	2.563	65	45395	4.564	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	91.200%	
20) 2-Butanone-d5	4.621	46	270385	72.999	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	146.000%#	
24) Chloroform-d	5.055	84	229165	5.332	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	106.600%	
26) 1,2-Dichloroethane-d4	5.695	65	108730	5.746	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	115.000%	
32) Benzene-d6	5.721	84	443444	5.042	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.800%	
36) 1,2-Dichloropropane-d6	6.685	67	140485	5.427	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	108.600%	
41) Toluene-d8	7.894	98	379543	4.710	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	94.200%	
43) trans-1,3-Dichloroprop...	8.177	79	48152	4.866	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	97.400%	
46) 2-Hexanone-d5	8.631	63	219395	67.643	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	135.280%#	
56) 1,1,2,2-Tetrachloroeth...	10.753	84	111925	6.057	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	121.200%#	
66) 1,2-Dichlorobenzene-d4	12.190	152	145851	5.888	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	117.800%	
Target Compounds					Qvalue	
13) Acetone	2.631	43	50178	18.369	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111023\
Data File : VU056192.D
Acq On : 11 Nov 2023 00:13
Operator : MD/SY
Sample : 05339-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCDJ2

Quant Time: Nov 11 01:44:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 11 00:40:59 2023
Response via : Initial Calibration

