



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111025\
Data File : VU063688.D
Acq On : 10 Nov 2025 08:26
Operator : MD/SY
Sample : BFB
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U111025W.M
Title : SW846 8260
Last Update : Tue Nov 11 01:41:19 2025



AutoFind: Scans 2898, 2899, 2900; Background Corrected with Scan 2885

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	8622	PASS
75	95	30	60	56.0	23747	PASS
95	95	100	100	100.0	42411	PASS
96	95	5	9	6.7	2842	PASS
173	174	0.00	2	0.6	186	PASS
174	95	50	100	77.1	32715	PASS
175	174	5	9	8.1	2642	PASS
176	174	95	101	97.1	31779	PASS
177	176	5	9	6.8	2173	PASS