

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
 Data File : VU045732.D
 Acq On : 11 Nov 2021 12:45
 Operator : SY/MD
 Sample : VU1111WBSD01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1111WBSD01

Quant Time: Nov 12 06:06:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	108782	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	189166	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	185940	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	94661	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	86973	53.386	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.780%	
35) Dibromofluoromethane	5.292	113	65786	52.089	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.180%	
50) Toluene-d8	7.903	98	246790	52.550	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.100%	
62) 4-Bromofluorobenzene	10.636	95	94806	50.343	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.680%	

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	1.385	85	26842	19.031	ug/l	99
3) Chloromethane	1.523	50	30046	18.619	ug/l	99
4) Vinyl Chloride	1.604	62	30435	19.361	ug/l	99
5) Bromomethane	1.851	94	19910	22.707	ug/l	98
6) Chloroethane	1.932	64	18472	18.006	ug/l	98
7) Trichlorofluoromethane	2.138	101	41652	20.037	ug/l	98
8) Diethyl Ether	2.375	74	17874	21.722	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.581	101	24981	20.580	ug/l	98
10) Methyl Iodide	2.723	142	26261	16.532	ug/l	92
11) Tert butyl alcohol	3.176	59	35708	110.737	ug/l	100
12) 1,1-Dichloroethene	2.581	96	22717	18.982	ug/l	96
13) Acrolein	2.488	56	4676	100.068	ug/l	94
14) Allyl chloride	2.925	41	41693	19.510	ug/l	99
15) Acrylonitrile	3.314	53	94203	109.420	ug/l	97
16) Acetone	2.623	43	82124	99.027	ug/l	100
17) Carbon Disulfide	2.793	76	57522	17.159	ug/l	100
18) Methyl Acetate	2.948	43	60776	21.516	ug/l	100
19) Methyl tert-butyl Ether	3.366	73	89666	20.814	ug/l	99
20) Methylene Chloride	3.047	84	29949	20.784	ug/l	97
21) trans-1,2-Dichloroethene	3.356	96	25501	20.481	ug/l	96
22) Diisopropyl ether	3.996	45	88265	21.036	ug/l	98
23) Vinyl Acetate	3.957	43	338850	104.999	ug/l	99
24) 1,1-Dichloroethane	3.871	63	52072	20.825	ug/l	97
25) 2-Butanone	4.703	43	122731	106.702	ug/l	99
26) 2,2-Dichloropropane	4.671	77	40495	19.544	ug/l	100
27) cis-1,2-Dichloroethene	4.671	96	30051	20.549	ug/l	99
28) Bromochloromethane	4.977	49	24964	20.942	ug/l	98
29) Tetrahydrofuran	5.051	42	74813	106.333	ug/l	99
30) Chloroform	5.092	83	53332	21.567	ug/l	98
31) Cyclohexane	5.391	56	45106	18.975	ug/l	95
32) 1,1,1-Trichloroethane	5.321	97	43579	20.665	ug/l	98
36) 1,1-Dichloropropene	5.530	75	36787	19.944	ug/l	99
37) Ethyl Acetate	4.806	43	46131	21.772	ug/l	99
38) Carbon Tetrachloride	5.530	117	36821	19.801	ug/l	100
39) Methylcyclohexane	6.768	83	42525	18.793	ug/l	97
40) Benzene	5.777	78	111045	20.334	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
 Data File : VU045732.D
 Acq On : 11 Nov 2021 12:45
 Operator : SY/MD
 Sample : VU1111WBSD01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1111WBSD01

Quant Time: Nov 12 06:06:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.973	41	23982	21.418	ug/l	95
42) 1,2-Dichloroethane	5.797	62	42854	21.725	ug/l	99
43) Isopropyl Acetate	5.912	43	67459	20.815	ug/l	99
44) Trichloroethene	6.546	130	27162	20.779	ug/l	99
45) 1,2-Dichloropropane	6.793	63	31153	21.560	ug/l	98
46) Dibromomethane	6.922	93	21580	21.554	ug/l	100
47) Bromodichloromethane	7.108	83	40746	21.449	ug/l	96
48) Methyl methacrylate	6.960	41	30902	20.717	ug/l	98
49) 1,4-Dioxane	6.964	88	16864	439.624	ug/l #	99
51) 4-Methyl-2-Pentanone	7.793	43	228955	107.097	ug/l	100
52) Toluene	7.973	92	70493	21.085	ug/l	98
53) t-1,3-Dichloropropene	8.214	75	42446	19.441	ug/l	99
54) cis-1,3-Dichloropropene	7.610	75	47886	20.650	ug/l	95
55) 1,1,2-Trichloroethane	8.404	97	30372	21.919	ug/l	98
56) Ethyl methacrylate	8.333	69	44218	19.353	ug/l	98
57) 1,3-Dichloropropane	8.578	76	53399	21.397	ug/l	100
58) 2-Chloroethyl Vinyl ether	7.465	63	97403	76.355	ug/l	98
59) 2-Hexanone	8.684	43	173904	98.205	ug/l	99
60) Dibromochloromethane	8.813	129	30798	21.345	ug/l	98
61) 1,2-Dibromoethane	8.925	107	31240	21.021	ug/l	100
64) Tetrachloroethene	8.555	164	23331	21.862	ug/l	98
65) Chlorobenzene	9.449	112	72545	19.863	ug/l	100
66) 1,1,1,2-Tetrachloroethane	9.539	131	26911	20.659	ug/l	98
67) Ethyl Benzene	9.575	91	130831	20.003	ug/l	99
68) m/p-Xylenes	9.697	106	102282	41.465	ug/l	98
69) o-Xylene	10.105	106	48863	20.375	ug/l	100
70) Styrene	10.118	104	83142	20.493	ug/l	98
71) Bromoform	10.295	173	21835	19.704	ug/l #	99
73) Isopropylbenzene	10.488	105	129944	21.009	ug/l	100
74) N-amyl acetate	10.320	43	51995	18.635	ug/l	98
75) 1,1,2,2-Tetrachloroethane	10.783	83	52111	21.265	ug/l	99
76) 1,2,3-Trichloropropane	10.825	75	56806	19.788	ug/l	92
77) Bromobenzene	10.787	156	31722	21.124	ug/l	95
78) n-propylbenzene	10.909	91	158944	20.648	ug/l	100
79) 2-Chlorotoluene	10.989	91	95031	20.895	ug/l	98
80) 1,3,5-Trimethylbenzene	11.092	105	112329	20.786	ug/l	99
81) trans-1,4-Dichloro-2-b...	10.549	75	13983	17.994	ug/l	95
82) 4-Chlorotoluene	11.099	91	110408	20.700	ug/l	99
83) tert-Butylbenzene	11.423	119	105293	20.440	ug/l	99
84) 1,2,4-Trimethylbenzene	11.472	105	111966	20.847	ug/l	100
85) sec-Butylbenzene	11.648	105	140155	20.515	ug/l	99
86) p-Isopropyltoluene	11.796	119	114022	19.972	ug/l	100
87) 1,3-Dichlorobenzene	11.748	146	60445	19.949	ug/l	99
88) 1,4-Dichlorobenzene	11.838	146	61038	19.918	ug/l	99
89) n-Butylbenzene	12.211	91	103247	18.774	ug/l	99
90) Hexachloroethane	12.478	117	19038	18.485	ug/l	98
91) 1,2-Dichlorobenzene	12.214	146	60325	20.464	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.999	75	11568	20.006	ug/l	99
93) 1,2,4-Trichlorobenzene	13.844	180	37757	18.524	ug/l	98
94) Hexachlorobutadiene	14.025	225	16250	18.470	ug/l	98
95) Naphthalene	14.089	128	139170	18.817	ug/l	99
96) 1,2,3-Trichlorobenzene	14.333	180	40290	19.901	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
Data File : VU045732.D
Acq On : 11 Nov 2021 12:45
Operator : SY/MD
Sample : VU1111WBSD01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01

Quant Time: Nov 12 06:06:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
Data File : VU045732.D
Acq On : 11 Nov 2021 12:45
Operator : SY/MD
Sample : VU1111WBSD01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01

Quant Time: Nov 12 06:06:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration





Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.375 min Scan# 11

Delta R.T. 0.000 min

Lab File: VU045732.D

Acq: 11 Nov 2021 12:45

Instrument :

MSVOA_U

ClientSampleId :





VU1111WBSD01

Tgt Ion:168 Resp: 108782

Ion Ratio Lower Upper

168 100

99 62.9 51.0 76.6



Dichlorodifluoromethane

Concen: 19.031 ug/l

RT: 1.385 min Scan# 14

Delta R.T. 0.000 min

Lab File: VU045732.D




Acq: 11 Nov 2021 12:45


Tgt Ion: 85 Resp: 26842

Ion Ratio Lower Upper

85 100

87 32.5 15.9 47.7









#3
Chloromethane
Concen: 18.619 ug/l
RT: 1.523 min Scan# 51
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion: 50 Resp: 30046
Ion Ratio Lower Upper
50 100
52 32.8 26.6 40.0



#4
Vinyl Chloride
Concen: 19.361 ug/l
RT: 1.604 min Scan# 82
Delta R.T. -0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 62 Resp: 30435
Ion Ratio Lower Upper
62 100
64 31.5 25.8 38.6









#5
Bromomethane
Concen: 22.707 ug/l
RT: 1.851 min Scan# 11
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01

Tgt Ion: 94 Resp: 19910
Ion Ratio Lower Upper
94 100
96 92.6 75.4 113.2



#6
Chloroethane
Concen: 18.006 ug/l
RT: 1.932 min Scan# 184
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 64 Resp: 18472
Ion Ratio Lower Upper
64 100
66 32.2 26.7 40.1









#7
Trichlorofluoromethane
Concen: 20.037 ug/l
RT: 2.138 min Scan# 248
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01

Tgt Ion:101 Resp: 41652
Ion Ratio Lower Upper
101 100
103 65.5 51.4 77.0




#8
Diethyl Ether
Concen: 21.722 ug/l
RT: 2.375 min Scan# 322
Delta R.T. -0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 74 Resp: 17874
Ion Ratio Lower Upper
74 100
45 90.1 47.3 141.8




Abundance Scan 387 (2.584 min): VU045552.D\data.ms (-37




m/z-->

Abundance Scan 386 (2.581 min): VU045732.D\data.ms




m/z-->

Abundance Scan 386 (2.581 min): VU045732.D\data.ms (-29




m/z-->

Abundance Scan 431 (2.726 min): VU045552.D\data.ms (-41




m/z-->

Abundance Scan 430 (2.723 min): VU045732.D\data.ms



m/z-->

Abundance Scan 430 (2.723 min): VU045732.D\data.ms (-35



m/z-->

#9

1,1,2-Trichlorotrifluoroethane

Concen: 20.580 ug/l

RT: 2.581 min Scan# 386

Delta R.T. -0.003 min

Lab File: VU045732.D

Acq: 11 Nov 2021 12:45

Instrument :

MSVOA_U

ClientSampleId :

VU1111WBSD01


Tgt Ion:101 Resp: 24981

Ion Ratio Lower Upper

101 100

85 43.3 35.0 52.4

151 70.9 54.6 82.0



#10

Methyl Iodide

Concen: 16.532 ug/l

RT: 2.723 min Scan# 430

Delta R.T. -0.003 min

Lab File: VU045732.D

Acq: 11 Nov 2021 12:45


Tgt Ion:142 Resp: 26261


Ion Ratio Lower Upper

142 100

127 46.5 32.2 48.2

141 13.4 11.3 16.9








#11
Tert butyl alcohol
Concen: 110.737 ug/l
RT: 3.176 min Scan# 511
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion: 59 Resp: 35708
Ion Ratio Lower Upper
59 100
57 10.4 8.3 12.5



#12
1,1-Dichloroethene
Concen: 18.982 ug/l
RT: 2.581 min Scan# 386
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 96 Resp: 22717
Ion Ratio Lower Upper
96 100
61 165.2 137.1 205.7
98 62.6 51.5 77.3









#13
Acrolein
Concen: 100.068 ug/l
RT: 2.488 min Scan# 357
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion: 56 Resp: 4676
Ion Ratio Lower Upper
56 100
55 75.0 56.3 84.5



#14
Allyl chloride
Concen: 19.510 ug/l
RT: 2.925 min Scan# 493
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 41 Resp: 41693
Ion Ratio Lower Upper
41 100
39 62.4 49.7 74.5
76 35.1 27.7 41.5









#15
Acrylonitrile
Concen: 109.420 ug/l
RT: 3.314 min Scan# 613
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion: 53 Resp: 94203
Ion Ratio Lower Upper
53 100
52 79.8 66.6 100.0
51 34.5 28.4 42.6



#16
Acetone
Concen: 99.027 ug/l
RT: 2.623 min Scan# 399
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 43 Resp: 82124
Ion Ratio Lower Upper
43 100
58 33.0 26.2 39.2









#17
Carbon Disulfide
Concen: 17.159 ug/l
RT: 2.793 min Scan# 41
Delta R.T. -0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion: 76 Resp: 57522
Ion Ratio Lower Upper
76 100
78 8.9 7.3 10.9



#18
Methyl Acetate
Concen: 21.516 ug/l
RT: 2.948 min Scan# 500
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 43 Resp: 60776
Ion Ratio Lower Upper
43 100
74 25.4 20.2 30.2









#19
Methyl tert-butyl Ether
Concen: 20.814 ug/l
RT: 3.366 min Scan# 61
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion: 73 Resp: 89666
Ion Ratio Lower Upper
73 100
57 22.2 18.1 27.1



#20
Methylene Chloride
Concen: 20.784 ug/l
RT: 3.047 min Scan# 531
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 84 Resp: 29949
Ion Ratio Lower Upper
84 100
49 117.0 96.4 144.6
51 34.7 30.5 45.7
86 61.2 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 20.481 ug/l
RT: 3.356 min Scan# 61
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion: 96 Resp: 25501
Ion Ratio Lower Upper
96 100
61 142.5 118.2 177.4
98 63.6 52.1 78.1



#22
Diisopropyl ether
Concen: 21.036 ug/l
RT: 3.996 min Scan# 826
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 45 Resp: 88265
Ion Ratio Lower Upper
45 100
43 73.2 60.0 90.0
87 28.6 22.3 33.5
59 12.7 9.9 14.9









#23
 Vinyl Acetate
 Concen: 104.999 ug/l
 RT: 3.957 min Scan# 814
 Delta R.T. 0.000 min
 Lab File: VU045732.D
 Acq: 11 Nov 2021 12:45




Instrument :
 MSVOA_U
 ClientSampleId :
 VU1111WBSD01

Tgt Ion: 43 Resp: 338850
 Ion Ratio Lower Upper
 43 100
 86 11.4 8.9 13.3



#24
 1,1-Dichloroethane
 Concen: 20.825 ug/l
 RT: 3.871 min Scan# 787
 Delta R.T. -0.003 min
 Lab File: VU045732.D
 Acq: 11 Nov 2021 12:45

Tgt Ion: 63 Resp: 52072
 Ion Ratio Lower Upper
 63 100
 98 6.2 3.5 10.6
 100 3.9 2.4 7.2









#25
2-Butanone
Concen: 106.702 ug/l
RT: 4.703 min Scan# 1046
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion: 43 Resp: 122731
Ion Ratio Lower Upper
43 100
72 27.8 22.0 33.0



#26
2,2-Dichloropropane
Concen: 19.544 ug/l
RT: 4.671 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 77 Resp: 40495
Ion Ratio Lower Upper
77 100
97 22.6 11.4 34.1









#27
cis-1,2-Dichloroethene
Concen: 20.549 ug/l
RT: 4.671 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01

Tgt Ion: 96 Resp: 30051




Ion	Ratio	Lower	Upper
96	100		
61	148.4	0.0	299.2
98	65.0	0.0	127.6




#28
Bromochloromethane
Concen: 20.942 ug/l
RT: 4.977 min Scan# 1131
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 49 Resp: 24964

Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	3.4
130	74.0	57.5	86.3









#29
Tetrahydrofuran
Concen: 106.333 ug/l
RT: 5.051 min Scan# 1154
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion:	42	Resp:	74813
Ion	Ratio	Lower	Upper
42	100		
72	48.4	38.2	57.4
71	44.8	35.4	53.2



#30
Chloroform
Concen: 21.567 ug/l
RT: 5.092 min Scan# 1167
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45


Tgt Ion:	83	Resp:	53332
Ion	Ratio	Lower	Upper
83	100		
85	66.8	52.0	78.0








#31
Cyclohexane
Concen: 18.975 ug/l
RT: 5.391 min Scan# 11
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01







Tgt Ion: 56 Resp: 45106
Ion Ratio Lower Upper
56 100
69 34.7 26.4 39.6
84 90.4 68.4 102.6



#32
1,1,1-Trichloroethane
Concen: 20.665 ug/l
RT: 5.321 min Scan# 1238
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 97 Resp: 43579
Ion Ratio Lower Upper
97 100
99 62.7 50.8 76.2
61 45.2 38.3 57.5





#33

1,2-Dichloroethane-d4

Concen: 53.386 ug/l

RT: 5.703 min Scan# 11

Delta R.T. 0.000 min

Lab File: VU045732.D

Acq: 11 Nov 2021 12:45

Instrument :

MSVOA_U

ClientSampleId :





VU1111WBSD01

Tgt Ion: 65 Resp: 86973

Ion Ratio Lower Upper

65 100

67 53.2 0.0 105.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.253 min Scan# 1528

Delta R.T. 0.000 min

Lab File: VU045732.D

Acq: 11 Nov 2021 12:45




Tgt Ion: 114 Resp: 189166

Ion Ratio Lower Upper

114 100

63 22.0 0.0 45.6

88 17.4 0.0 33.8









#35
Dibromofluoromethane
Concen: 52.089 ug/l
RT: 5.292 min Scan# 11
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion:113 Resp: 65786
Ion Ratio Lower Upper
113 100
111 104.5 82.8 124.2
192 18.4 14.8 22.2



#36
1,1-Dichloropropene
Concen: 19.944 ug/l
RT: 5.530 min Scan# 1303
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 75 Resp: 36787
Ion Ratio Lower Upper
75 100
110 32.7 16.7 50.0
77 30.9 25.0 37.4









#37
Ethyl Acetate
Concen: 21.772 ug/l
RT: 4.806 min Scan# 1078
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01

Tgt Ion: 43	Resp: 46131
Ion Ratio	Lower Upper
43	100
61	15.4 11.6 17.4
70	11.3 9.1 13.7



#38
Carbon Tetrachloride
Concen: 19.801 ug/l
RT: 5.530 min Scan# 1303
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 117	Resp: 36821
Ion Ratio	Lower Upper
117	100
119	96.7 77.4 116.0
121	31.0 24.8 37.2





#39

Methylcyclohexane

Concen: 18.793 ug/l

RT: 6.768 min Scan# 1

Delta R.T. 0.000 min

Lab File: VU045732.D

Acq: 11 Nov 2021 12:45

Instrument :

MSVOA_U

ClientSampleId :

VU1111WBSD01





Tgt Ion: 83 Resp: 42525

Ion Ratio Lower Upper

83 100

55 76.7 63.6 95.4

98 48.9 37.1 55.7



#40

Benzene

Concen: 20.334 ug/l

RT: 5.777 min Scan# 1380

Delta R.T. 0.000 min

Lab File: VU045732.D




Acq: 11 Nov 2021 12:45

Tgt Ion: 78 Resp: 111045

Ion Ratio Lower Upper

78 100

77 23.7 19.0 28.4









#41
Methacrylonitrile
Concen: 21.418 ug/l
RT: 4.973 min Scan# 1129
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01

Tgt Ion: 41 Resp: 23982




Ion	Ratio	Lower	Upper
41	100		
39	52.2	43.4	65.0
67	81.4	61.0	91.4
52	31.3	23.4	35.0




#42
1,2-Dichloroethane
Concen: 21.725 ug/l
RT: 5.797 min Scan# 1386
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 62 Resp: 42854

Ion	Ratio	Lower	Upper
62	100		
98	8.9	0.0	18.4









#43
Isopropyl Acetate
Concen: 20.815 ug/l
RT: 5.912 min Scan# 1422
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion: 43	Resp: 67459
Ion Ratio	Lower Upper
43	100
61	21.6 16.9 25.3
87	13.7 11.0 16.4



#44
Trichloroethene
Concen: 20.779 ug/l
RT: 6.546 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 130	Resp: 27162
Ion Ratio	Lower Upper
130	100
95	106.3 0.0 211.4





#45

1,2-Dichloropropane

Concen: 21.560 ug/l

RT: 6.793 min Scan# 1

Delta R.T. 0.000 min

Lab File: VU045732.D

Acq: 11 Nov 2021 12:45

Instrument :

MSVOA_U

ClientSampleId :





VU1111WBSD01

Tgt Ion: 63 Resp: 31153

Ion Ratio Lower Upper

63 100

65 31.1 25.7 38.5



#46

Dibromomethane

Concen: 21.554 ug/l

RT: 6.922 min Scan# 1736

Delta R.T. 0.003 min

Lab File: VU045732.D

Acq: 11 Nov 2021 12:45




Tgt Ion: 93 Resp: 21580

Ion Ratio Lower Upper

93 100

95 82.7 66.2 99.2

174 88.3 71.0 106.6









#47
Bromodichloromethane
Concen: 21.449 ug/l
RT: 7.108 min Scan# 11
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion: 83 Resp: 40746
Ion Ratio Lower Upper
83 100
85 65.7 49.8 74.6
127 9.0 6.2 9.2



#48
Methyl methacrylate
Concen: 20.717 ug/l
RT: 6.960 min Scan# 1748
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 41 Resp: 30902
Ion Ratio Lower Upper
41 100
69 90.3 70.2 105.2
39 54.2 43.9 65.9









#49
1,4-Dioxane
Concen: 439.624 ug/l
RT: 6.964 min Scan# 11
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01

Tgt Ion: 88 Resp: 16864
Ion Ratio Lower Upper
88 100
43 29.9 0.0 0.0#
58 72.2 58.7 88.1



#50
Toluene-d8
Concen: 52.550 ug/l
RT: 7.903 min Scan# 2041
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 98 Resp: 246790
Ion Ratio Lower Upper
98 100
100 65.8 52.5 78.7





#51

4-Methyl-2-Pentanone

Concen: 107.097 ug/l

RT: 7.793 min Scan# 2007

Delta R.T. 0.000 min

Lab File: VU045732.D

Acq: 11 Nov 2021 12:45

Instrument :

MSVOA_U

ClientSampleId :





VU1111WBSD01

Tgt Ion: 43 Resp: 228955

Ion Ratio Lower Upper

43 100

58 42.3 33.8 50.8



#52

Toluene

Concen: 21.085 ug/l

RT: 7.973 min Scan# 2063

Delta R.T. 0.003 min

Lab File: VU045732.D




Acq: 11 Nov 2021 12:45


Tgt Ion: 92 Resp: 70493

Ion Ratio Lower Upper

92 100

91 167.3 135.6 203.4





#53

t-1,3-Dichloropropene

Concen: 19.441 ug/l

RT: 8.214 min Scan# 2137

Delta R.T. 0.003 min

Lab File: VU045732.D

Acq: 11 Nov 2021 12:45

Instrument :

MSVOA_U

ClientSampleId :





VU1111WBSD01

Tgt Ion: 75 Resp: 42446

Ion Ratio Lower Upper

75 100

77 31.9 25.8 38.8



#54

cis-1,3-Dichloropropene

Concen: 20.650 ug/l

RT: 7.610 min Scan# 1950

Delta R.T. 0.000 min

Lab File: VU045732.D

Acq: 11 Nov 2021 12:45




Tgt Ion: 75 Resp: 47886


Ion Ratio Lower Upper

75 100

77 33.2 24.6 37.0

39 43.3 37.1 55.7









#55
1,1,2-Trichloroethane
Concen: 21.919 ug/l
RT: 8.404 min Scan# 2196
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01

Tgt Ion: 97 Resp: 30372




Ion	Ratio	Lower	Upper
97	100		
83	86.2	70.3	105.5
85	58.3	45.5	68.3
99	63.8	49.7	74.5




#56
Ethyl methacrylate
Concen: 19.353 ug/l
RT: 8.333 min Scan# 2175
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 69 Resp: 44218

Ion	Ratio	Lower	Upper
69	100		
41	61.7	51.0	76.4
39	33.0	26.6	39.8









#57
1,3-Dichloropropane
Concen: 21.397 ug/l
RT: 8.578 min Scan# 21
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01

Tgt Ion: 76 Resp: 53399
Ion Ratio Lower Upper
76 100
78 32.3 25.8 38.8



#58
2-Chloroethyl Vinyl ether
Concen: 76.355 ug/l
RT: 7.465 min Scan# 1905
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 63 Resp: 97403
Ion Ratio Lower Upper
63 100
106 27.0 21.0 31.4









#59
2-Hexanone
Concen: 98.205 ug/l
RT: 8.684 min Scan# 21
Delta R.T. -0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion: 43 Resp: 173904
Ion Ratio Lower Upper
43 100
58 58.2 29.5 88.6



#60
Dibromochloromethane
Concen: 21.345 ug/l
RT: 8.813 min Scan# 2324
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 129 Resp: 30798
Ion Ratio Lower Upper
129 100
127 75.5 38.5 115.3









#61
1,2-Dibromoethane
Concen: 21.021 ug/l
RT: 8.925 min Scan# 21
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01

Tgt Ion:107 Resp: 31240
Ion Ratio Lower Upper
107 100
109 95.0 75.8 113.8



#62
4-Bromofluorobenzene
Concen: 50.343 ug/l
RT: 10.636 min Scan# 2891
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 95 Resp: 94806
Ion Ratio Lower Upper
95 100
174 73.3 0.0 141.8
176 71.6 0.0 138.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.420 min Scan# 2117
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion:117 Resp: 185940
Ion Ratio Lower Upper
117 100
82 58.8 47.8 71.8
119 31.6 25.8 38.8



#64
Tetrachloroethene
Concen: 21.862 ug/l
RT: 8.555 min Scan# 2244
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion:164 Resp: 23331
Ion Ratio Lower Upper
164 100
166 126.5 102.2 153.4
129 96.1 78.2 117.4
131 97.2 75.1 112.7









#65
Chlorobenzene
Concen: 19.863 ug/l
RT: 9.449 min Scan# 21
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion:112 Resp: 72545
Ion Ratio Lower Upper
112 100
114 32.4 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 20.659 ug/l
RT: 9.539 min Scan# 2550
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion:131 Resp: 26911
Ion Ratio Lower Upper
131 100
133 94.9 47.9 143.6
119 66.5 34.5 103.6





#67

Ethyl Benzene

Concen: 20.003 ug/l

RT: 9.575 min Scan# 2560

Delta R.T. 0.003 min

Lab File: VU045732.D

Acq: 11 Nov 2021 12:45

Instrument :

MSVOA_U

ClientSampleId :





VU1111WBSD01

Tgt Ion: 91 Resp: 130831

Ion Ratio Lower Upper

91 100

106 30.3 24.6 36.8



#68

m/p-Xylenes

Concen: 41.465 ug/l

RT: 9.697 min Scan# 2599

Delta R.T. 0.000 min

Lab File: VU045732.D




Acq: 11 Nov 2021 12:45


Tgt Ion: 106 Resp: 102282

Ion Ratio Lower Upper

106 100

91 202.9 164.7 247.1









#69
o-Xylene
Concen: 20.375 ug/l
RT: 10.105 min Scan# 21
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01

Tgt Ion:106 Resp: 48863
Ion Ratio Lower Upper
106 100
91 217.7 108.7 326.3



#70
Styrene
Concen: 20.493 ug/l
RT: 10.118 min Scan# 2730
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion:104 Resp: 83142
Ion Ratio Lower Upper
104 100
78 50.2 41.8 62.6
103 53.7 43.8 65.6









#71
Bromoform
Concen: 19.704 ug/l
RT: 10.295 min Scan# 21
Delta R.T. 0.004 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion:173 Resp: 21835
Ion Ratio Lower Upper
173 100
175 48.9 24.1 72.2
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 11.816 min Scan# 3258
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion:152 Resp: 94661
Ion Ratio Lower Upper
152 100
115 71.7 42.7 128.1
150 162.9 0.0 346.8









#73
Isopropylbenzene
Concen: 21.009 ug/l
RT: 10.488 min Scan# 2844
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion:105 Resp: 129944
Ion Ratio Lower Upper
105 100
120 25.6 12.9 38.6



#74
N-amy1 acetate
Concen: 18.635 ug/l
RT: 10.320 min Scan# 2793
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 43 Resp: 51995
Ion Ratio Lower Upper
43 100
70 47.1 36.9 55.3
55 28.3 20.6 31.0
61 25.4 20.5 30.7









#75
1,1,2,2-Tetrachloroethane
Concen: 21.265 ug/l
RT: 10.783 min Scan# 2937
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion: 83 Resp: 52111
Ion Ratio Lower Upper
83 100
131 9.3 4.4 13.2
85 64.7 31.9 95.9



#76
1,2,3-Trichloropropane
Concen: 19.788 ug/l
RT: 10.825 min Scan# 2950
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 75 Resp: 56806
Ion Ratio Lower Upper
75 100
77 29.7 17.1 51.3









#77
Bromobenzene
Concen: 21.124 ug/l
RT: 10.787 min Scan# 2937
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01

Tgt Ion:156 Resp: 31722
Ion Ratio Lower Upper
156 100
77 171.4 91.1 273.4
158 95.8 47.9 143.6



#78
n-propylbenzene
Concen: 20.648 ug/l
RT: 10.909 min Scan# 2976
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 91 Resp: 158944
Ion Ratio Lower Upper
91 100
120 21.6 10.9 32.7









#79
2-Chlorotoluene
Concen: 20.895 ug/l
RT: 10.989 min Scan# 3000
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01

Tgt Ion: 91 Resp: 95031
Ion Ratio Lower Upper
91 100
126 32.2 15.6 46.8



#80
1,3,5-Trimethylbenzene
Concen: 20.786 ug/l
RT: 11.092 min Scan# 3033
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 105 Resp: 112329
Ion Ratio Lower Upper
105 100
120 47.3 23.9 71.8









#81
trans-1,4-Dichloro-2-butene
Concen: 17.994 ug/l
RT: 10.549 min Scan# 2864
Delta R.T. -0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion:	75	Resp:	13983
Ion	Ratio	Lower	Upper
75	100		
53	100.0	76.2	114.4
89	46.6	34.0	51.0



#82
4-Chlorotoluene
Concen: 20.700 ug/l
RT: 11.099 min Scan# 3035
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion:	91	Resp:	110408
Ion	Ratio	Lower	Upper
91	100		
126	28.1	13.8	41.4









#83
tert-Butylbenzene
Concen: 20.440 ug/l
RT: 11.423 min Scan# 3136
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion:119 Resp: 105293
Ion Ratio Lower Upper
119 100
91 59.6 30.5 91.5
134 23.0 11.5 34.5



#84
1,2,4-Trimethylbenzene
Concen: 20.847 ug/l
RT: 11.472 min Scan# 3151
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion:105 Resp: 111966
Ion Ratio Lower Upper
105 100
120 44.3 22.2 66.6









#85
 sec-Butylbenzene
 Concen: 20.515 ug/l
 RT: 11.648 min Scan# 31
 Delta R.T. 0.003 min
 Lab File: VU045732.D
 Acq: 11 Nov 2021 12:45




Instrument :
 MSVOA_U
 ClientSampleId :
 VU1111WBSD01


Tgt Ion:105 Resp: 140155
 Ion Ratio Lower Upper
 105 100
 134 18.6 9.4 28.3



#86
 p-Isopropyltoluene
 Concen: 19.972 ug/l
 RT: 11.796 min Scan# 3252
 Delta R.T. 0.000 min
 Lab File: VU045732.D
 Acq: 11 Nov 2021 12:45

Tgt Ion:119 Resp: 114022
 Ion Ratio Lower Upper
 119 100
 134 25.2 12.7 38.1
 91 24.5 12.2 36.4





#87

1,3-Dichlorobenzene

Concen: 19.949 ug/l

RT: 11.748 min Scan# 3237

Delta R.T. 0.000 min

Lab File: VU045732.D

Acq: 11 Nov 2021 12:45

Instrument :

MSVOA_U

ClientSampleId :

VU1111WBSD01





Tgt Ion:146 Resp: 60445

Ion Ratio Lower Upper

146 100

111 42.4 21.3 63.9

148 62.3 31.7 95.1



#88

1,4-Dichlorobenzene

Concen: 19.918 ug/l

RT: 11.838 min Scan# 3265

Delta R.T. 0.000 min

Lab File: VU045732.D

Acq: 11 Nov 2021 12:45




Tgt Ion:146 Resp: 61038

Ion Ratio Lower Upper

146 100

111 43.0 20.6 61.9

148 64.0 31.9 95.7









#89
n-Butylbenzene
Concen: 18.774 ug/l
RT: 12.211 min Scan# 311
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01

Tgt Ion: 91 Resp: 103247
Ion Ratio Lower Upper
91 100
92 53.6 27.1 81.2
134 24.5 11.8 35.4



#90
Hexachloroethane
Concen: 18.485 ug/l
RT: 12.478 min Scan# 3464
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 117 Resp: 19038
Ion Ratio Lower Upper
117 100
201 66.3 33.8 101.4









#91
1,2-Dichlorobenzene
Concen: 20.464 ug/l
RT: 12.214 min Scan# 311
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion:146 Resp: 60325
Ion Ratio Lower Upper
146 100
111 44.0 21.9 65.5
148 64.3 32.1 96.3



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.006 ug/l
RT: 12.999 min Scan# 3626
Delta R.T. 0.000 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion: 75 Resp: 11568
Ion Ratio Lower Upper
75 100
155 75.7 37.8 113.3
157 98.9 48.4 145.2









#93
1,2,4-Trichlorobenzene
Concen: 18.524 ug/l
RT: 13.844 min Scan# 3888
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45




Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01


Tgt Ion:180 Resp: 37757
Ion Ratio Lower Upper
180 100
182 95.3 47.0 141.2
145 33.2 15.6 46.7



#94
Hexachlorobutadiene
Concen: 18.470 ug/l
RT: 14.025 min Scan# 3945
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion:225 Resp: 16250
Ion Ratio Lower Upper
225 100
223 62.3 32.0 96.2
227 63.0 31.1 93.5









#95
Naphthalene
Concen: 18.817 ug/l
RT: 14.089 min Scan# 3964
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Instrument :
MSVOA_U
ClientSampleId :
VU1111WBSD01

Tgt Ion:128 Resp: 139170
Ion Ratio Lower Upper
128 100
127 12.8 10.6 15.8
129 10.7 8.6 13.0



#96
1,2,3-Trichlorobenzene
Concen: 19.901 ug/l
RT: 14.333 min Scan# 4041
Delta R.T. 0.003 min
Lab File: VU045732.D
Acq: 11 Nov 2021 12:45

Tgt Ion:180 Resp: 40290
Ion Ratio Lower Upper
180 100
182 93.2 47.7 143.1
145 34.2 16.7 50.0

