

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
 Data File : VU045738.D
 Acq On : 11 Nov 2021 15:04
 Operator : SY/MD
 Sample : M4606-18
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB-01-20211109

Quant Time: Nov 12 06:09:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	117426	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	217890	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	221835	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	106988	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	95896	54.530	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.060%
35) Dibromofluoromethane	5.292	113	75305	51.765	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.540%
50) Toluene-d8	7.902	98	288115	53.262	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.520%
62) 4-Bromofluorobenzene	10.635	95	106794	49.233	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.460%
Target Compounds						
16) Acetone	2.626	43	1484	1.658	ug/l #	76
44) Trichloroethene	6.546	130	1361	0.904	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
Data File : VU045738.D
Acq On : 11 Nov 2021 15:04
Operator : SY/MD
Sample : M4606-18
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB-01-20211109

Quant Time: Nov 12 06:09:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

