

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
 Data File : VU045739.D
 Acq On : 11 Nov 2021 15:27
 Operator : SY/MD
 Sample : M4606-22
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FB-01-20211109

Quant Time: Nov 12 06:09:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	119154	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	220085	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	226087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	110264	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	98411	55.149	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.300%
35) Dibromofluoromethane	5.292	113	76672	52.179	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.360%
50) Toluene-d8	7.899	98	290563	53.179	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.360%
62) 4-Bromofluorobenzene	10.636	95	108981	49.740	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.480%
Target Compounds						
16) Acetone	2.630	43	3389	3.731	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
Data File : VU045739.D
Acq On : 11 Nov 2021 15:27
Operator : SY/MD
Sample : M4606-22
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FB-01-20211109

Quant Time: Nov 12 06:09:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

