

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
 Data File : VU045740.D
 Acq On : 11 Nov 2021 15:50
 Operator : SY/MD
 Sample : M4606-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MH-117N-20211109

Quant Time: Nov 12 06:09:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	117035	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	216607	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	220854	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	109246	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	96205	54.889	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.780%	
35) Dibromofluoromethane	5.292	113	74793	51.718	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.440%	
50) Toluene-d8	7.899	98	283489	52.717	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.440%	
62) 4-Bromofluorobenzene	10.636	95	106608	49.438	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.880%	
Target Compounds						
16) Acetone	2.629	43	1378	1.544	ug/l	96
44) Trichloroethene	6.542	130	494m	0.330	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
Data File : VU045740.D
Acq On : 11 Nov 2021 15:50
Operator : SY/MD
Sample : M4606-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MH-117N-20211109

Quant Time: Nov 12 06:09:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

