

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
 Data File : VU045741.D
 Acq On : 11 Nov 2021 16:13
 Operator : SY/MD
 Sample : M4606-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MH-117T-20211109

Quant Time: Nov 12 06:10:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	117537	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	219057	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	228123	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	111404	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	96481	54.811	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.620%	
35) Dibromofluoromethane	5.292	113	76180	52.088	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.180%	
50) Toluene-d8	7.902	98	290548	53.425	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.860%	
62) 4-Bromofluorobenzene	10.635	95	109052	50.006	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.020%	
Target Compounds						
						Qvalue
16) Acetone	2.629	43	1510	1.685	ug/l	100
44) Trichloroethene	6.545	130	11196	7.396	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
Data File : VU045741.D
Acq On : 11 Nov 2021 16:13
Operator : SY/MD
Sample : M4606-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MH-117T-20211109

Quant Time: Nov 12 06:10:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

