

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
 Data File : VU045744.D
 Acq On : 11 Nov 2021 17:23
 Operator : SY/MD
 Sample : M4606-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MH-121.5N-20211109

Quant Time: Nov 12 06:11:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	116875	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	214099	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	225474	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	107937	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	97111	55.481	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.960%
35) Dibromofluoromethane	5.292	113	75891	53.092	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.180%
50) Toluene-d8	7.899	98	289792	54.520	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.040%
62) 4-Bromofluorobenzene	10.636	95	106306	49.876	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.760%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	3.182	59	108	0.312	ug/l #	72
16) Acetone	2.629	43	1283	1.440	ug/l #	76
44) Trichloroethene	6.549	130	477	0.322	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
Data File : VU045744.D
Acq On : 11 Nov 2021 17:23
Operator : SY/MD
Sample : M4606-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MH-121.5N-20211109

Quant Time: Nov 12 06:11:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

