

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
 Data File : VU045747.D
 Acq On : 11 Nov 2021 18:32
 Operator : SY/MD
 Sample : M4606-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MH-140B-20211109

Quant Time: Nov 12 06:12:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	115325	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	210670	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	222956	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	110583	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	95609	55.357	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 110.720%	
35) Dibromofluoromethane	5.292	113	75197	53.463	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.920%	
50) Toluene-d8	7.902	98	283190	54.145	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.300%	
62) 4-Bromofluorobenzene	10.636	95	106516	50.788	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.580%	
Target Compounds						
						Qvalue
16) Acetone	2.626	43	2169	2.467	ug/l	91
27) cis-1,2-Dichloroethene	4.674	96	819	0.528	ug/l	92
44) Trichloroethene	6.546	130	933	0.641	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
Data File : VU045747.D
Acq On : 11 Nov 2021 18:32
Operator : SY/MD
Sample : M4606-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MH-140B-20211109

Quant Time: Nov 12 06:12:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

