

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
 Data File : VU045749.D
 Acq On : 11 Nov 2021 19:18
 Operator : SY/MD
 Sample : M4606-13
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SPRING-20211109

Quant Time: Nov 12 06:12:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	118226	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	218799	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	229021	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	114347	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	99951	56.451	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.900%
35) Dibromofluoromethane	5.292	113	77413	52.993	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.980%
50) Toluene-d8	7.903	98	292282	53.808	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.620%
62) 4-Bromofluorobenzene	10.636	95	110354	50.663	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.320%
Target Compounds						
16) Acetone	2.629	43	1409	1.563	ug/l	Qvalue # 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
Data File : VU045749.D
Acq On : 11 Nov 2021 19:18
Operator : SY/MD
Sample : M4606-13
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SPRING-20211109

Quant Time: Nov 12 06:12:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

