

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
 Data File : VU045751.D
 Acq On : 11 Nov 2021 20:05
 Operator : SY/MD
 Sample : M4657-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TRIP-BLANK

Quant Time: Nov 12 06:13:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	113736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	215210	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	217673	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	107276	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	93308	54.780	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.560%	
35) Dibromofluoromethane	5.292	113	73923	51.448	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.900%	
50) Toluene-d8	7.903	98	279330	52.281	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.560%	
62) 4-Bromofluorobenzene	10.636	95	104427	48.741	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.480%	
Target Compounds						
16) Acetone	2.626	43	1453	1.676	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111121\
Data File : VU045751.D
Acq On : 11 Nov 2021 20:05
Operator : SY/MD
Sample : M4657-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 12 06:13:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

