

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111122\
 Data File : VU051819.D
 Acq On : 11 Nov 2022 16:49
 Operator : JC/MD
 Sample : N5603-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AH3

Quant Time: Nov 12 01:23:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 12 01:22:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	207463	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	209731	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	83462	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	61485	3.935	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.800%
7) Chloroethane-d5	1.909	69	61891	3.798	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	76.000%
11) 1,1-Dichloroethene-d2	2.565	65	26008	4.586	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.800%
20) 2-Butanone-d5	4.620	46	189296	33.452	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	66.900%
24) Chloroform-d	5.060	84	144800	3.960	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.200%
26) 1,2-Dichloroethane-d4	5.700	65	74773	3.786	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.800%
32) Benzene-d6	5.726	84	281205	4.249	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.000%
36) 1,2-Dichloropropane-d6	6.690	67	96016	4.149	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.000%
41) Toluene-d8	7.896	98	220287	4.232	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.600%
43) trans-1,3-Dichloroprop...	8.179	79	29552	4.176	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.600%
46) 2-Hexanone-d5	8.632	63	144045	33.815	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	67.620%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	83929	3.957	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.200%
66) 1,2-Dichlorobenzene-d4	12.192	152	84627	5.524	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111122\
Data File : VU051819.D
Acq On : 11 Nov 2022 16:49
Operator : JC/MD
Sample : N5603-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AH3

Quant Time: Nov 12 01:23:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 12 01:22:06 2022
Response via : Initial Calibration

