

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111124\
 Data File : VU061586.D
 Acq On : 11 Nov 2024 16:40
 Operator : MD/SY
 Sample : P4748-14
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCP68

Quant Time: Nov 12 03:46:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 12 03:43:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	175437	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	163908	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	70184	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	40657	3.412	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.200%
7) Chloroethane-d5	1.911	69	45277	4.144	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.800%
11) 1,1-Dichloroethene-d2	2.560	65	19283	3.782	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.600%
20) 2-Butanone-d5	4.624	46	114953	47.509	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.020%
24) Chloroform-d	5.052	84	97592	4.314	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.200%
26) 1,2-Dichloroethane-d4	5.692	65	51117	4.534	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.600%
32) Benzene-d6	5.718	84	199712	4.324	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.400%
36) 1,2-Dichloropropane-d6	6.679	67	62931	4.532	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.600%
41) Toluene-d8	7.891	98	176926	4.097	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.000%
43) trans-1,3-Dichloroprop...	8.174	79	19637	3.465	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	69.200%
46) 2-Hexanone-d5	8.627	63	100738	43.395	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.780%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	47275	4.502	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	58497	4.707	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.200%
Target Compounds						
13) Acetone	2.644	43	15844	10.339	ug/L	95
31) Carbon tetrachloride	5.515	117	17721	1.032	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111124\
Data File : VU061586.D
Acq On : 11 Nov 2024 16:40
Operator : MD/SY
Sample : P4748-14
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCP68

Quant Time: Nov 12 03:46:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 12 03:43:56 2024
Response via : Initial Calibration

