

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111124\
 Data File : VU061587.D
 Acq On : 11 Nov 2024 17:05
 Operator : MD/SY
 Sample : P4748-15
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCP69

Quant Time: Nov 12 03:47:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 12 03:43:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	187941	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	177727	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	75490	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	41256	3.232	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	64.600%
7) Chloroethane-d5	1.907	69	46113	3.940	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.800%
11) 1,1-Dichloroethene-d2	2.560	65	20216	3.701	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.000%
20) 2-Butanone-d5	4.624	46	116794	45.058	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.120%
24) Chloroform-d	5.052	84	98973	4.084	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.600%
26) 1,2-Dichloroethane-d4	5.695	65	52491	4.346	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.000%
32) Benzene-d6	5.717	84	208552	4.165	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.200%
36) 1,2-Dichloropropane-d6	6.682	67	64218	4.265	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.400%
41) Toluene-d8	7.891	98	182621	3.900	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.000%
43) trans-1,3-Dichloroprop...	8.174	79	19913	3.240	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	64.800%
46) 2-Hexanone-d5	8.627	63	101150	40.184	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.360%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	50135	4.403	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	60531	4.528	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111124\
Data File : VU061587.D
Acq On : 11 Nov 2024 17:05
Operator : MD/SY
Sample : P4748-15
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCP69

Quant Time: Nov 12 03:47:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 12 03:43:56 2024
Response via : Initial Calibration

