

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU111218\
 Data File : VU028060.D
 Acq On : 12 Nov 2018 13:20
 Operator : MD/SY
 Sample : J5885-09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-8

Quant Time: Nov 13 12:12:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	170584	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	324302	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	290389	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	111190	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	148664	58.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.24%	
35) Dibromofluoromethane	4.88	113	109263	53.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.88%	
50) Toluene-d8	7.57	98	421800	53.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.80%	
62) 4-Bromofluorobenzene	10.31	95	136648	46.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.80%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.23	96	6331	2.736	ug/l	95
30) Chloroform	4.68	83	4059	0.985	ug/l #	81
44) Trichloroethene	6.20	130	3034	1.359	ug/l	81
64) Tetrachloroethene	8.23	164	130099	69.093	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU111218\
Data File : VU028060.D
Acq On : 12 Nov 2018 13:20
Operator : MD/SY
Sample : J5885-09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-8

Quant Time: Nov 13 12:12:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

