

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU111218\
 Data File : VU028063.D
 Acq On : 12 Nov 2018 14:33
 Operator : MD/SY
 Sample : J5885-01DL 40X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RW-2DL

Quant Time: Nov 13 03:13:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	169775	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	322037	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	277241	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	95821	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	147513	58.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.88%	
35) Dibromofluoromethane	4.88	113	108113	53.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.48%	
50) Toluene-d8	7.57	98	415850	53.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.04%	
62) 4-Bromofluorobenzene	10.31	95	125080	43.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.46%	
Target Compounds						
64) Tetrachloroethene	8.23	164	34045	18.938	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU111218\
Data File : VU028063.D
Acq On : 12 Nov 2018 14:33
Operator : MD/SY
Sample : J5885-01DL 40X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RW-2DL

Quant Time: Nov 13 03:13:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

