

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU111218\
 Data File : VU028070.D
 Acq On : 12 Nov 2018 17:21
 Operator : MD/SY
 Sample : J5269-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SB-12-D

Quant Time: Nov 13 03:32:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	174832	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	330706	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	302907	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	118421	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	151909	58.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.88%	
35) Dibromofluoromethane	4.88	113	109522	52.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.04%	
50) Toluene-d8	7.57	98	438322	54.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.84%	
62) 4-Bromofluorobenzene	10.31	95	143054	48.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.30%	
Target Compounds						
16) Acetone	2.32	43	11970	7.422	ug/l	97
20) Methylene Chloride	2.70	84	13449	5.712	ug/l	89
43) Isopropyl Acetate	5.55	43	32900	4.138	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU111218\
Data File : VU028070.D
Acq On : 12 Nov 2018 17:21
Operator : MD/SY
Sample : J5269-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SB-12-D

Quant Time: Nov 13 03:32:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

