

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU111218\
 Data File : VU028077.D
 Acq On : 12 Nov 2018 20:10
 Operator : MD/SY
 Sample : J5887-16
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 W-3-S

Quant Time: Nov 13 04:01:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	175449	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	336937	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	300792	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	117307	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	153474	58.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.68%	
35) Dibromofluoromethane	4.88	113	109800	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	
50) Toluene-d8	7.57	98	434617	52.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.92%	
62) 4-Bromofluorobenzene	10.31	95	141367	46.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.40%	
Target Compounds						
16) Acetone	2.32	43	8947	5.528	ug/l	93
18) Methyl Acetate	2.62	43	9279	2.084	ug/l	98
20) Methylene Chloride	2.70	84	265161	112.215	ug/l	97
43) Isopropyl Acetate	5.55	43	205689	25.391	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU111218\
Data File : VU028077.D
Acq On : 12 Nov 2018 20:10
Operator : MD/SY
Sample : J5887-16
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
W-3-S

Quant Time: Nov 13 04:01:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

