

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111218\
 Data File : VU028055.D
 Acq On : 12 Nov 2018 11:01
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 13 02:53:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.635	0.523	17.6	70	0.00
3 P	Chloromethane	0.768	0.802	-4.4	94	0.00
4 C	Vinyl Chloride	0.763	0.732	4.1#	87	0.00
5 T	Bromomethane	0.336	0.342	-1.8	105	0.00
6 T	Chloroethane	0.439	0.465	-5.9	96	0.00
7 T	Trichlorofluoromethane	0.881	0.906	-2.8	92	0.00
8 T	Diethyl Ether	0.377	0.394	-4.5	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.524	0.544	-3.8	93	0.00
10 T	Methyl Iodide	0.392	0.542	-38.3#	117	0.00
11 T	Tert butyl alcohol	0.197	0.160	18.8	75	0.00
12 CM	1,1-Dichloroethene	0.524	0.495	5.5#	87	0.00
13 T	Acrolein	0.078	0.093	-19.2	126	0.00
14 T	Allyl chloride	1.281	1.248	2.6	92	0.00
15 T	Acrylonitrile	0.483	0.476	1.4	87	0.00
16 T	Acetone	0.461	0.462	-0.2	96	0.00
17 T	Carbon Disulfide	1.713	1.670	2.5	89	0.00
18 T	Methyl Acetate	1.269	1.243	2.0	93	0.00
19 T	Methyl tert-butyl Ether	2.113	2.166	-2.5	93	0.00
20 T	Methylene Chloride	0.673	0.675	-0.3	92	0.00
21 T	trans-1,2-Dichloroethene	0.605	0.583	3.6	89	0.00
22 T	Diisopropyl ether	2.644	2.751	-4.0	94	0.00
23 T	Vinyl Acetate	2.272	2.368	-4.2	92	0.00
24 P	1,1-Dichloroethane	1.317	1.346	-2.2	94	0.00
25 T	2-Butanone	0.734	0.720	1.9	87	0.00
26 T	2,2-Dichloropropane	0.988	1.075	-8.8	103	0.00
27 T	cis-1,2-Dichloroethene	0.678	0.710	-4.7	92	0.00
28 T	Bromochloromethane	0.617	0.703	-13.9	110	0.00
29 T	Tetrahydrofuran	0.490	0.444	9.4	82	0.00
30 C	Chloroform	1.208	1.212	-0.3#	93	0.00
31 T	Cyclohexane	1.547	1.290	16.6	93	0.00
32 T	1,1,1-Trichloroethane	0.961	1.014	-5.5	94	0.00
33 S	1,2-Dichloroethane-d4	0.743	0.784	-5.5	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.315	0.334	-6.0	101	0.00
36 T	1,1-Dichloropropene	0.524	0.560	-6.9	96	0.00
37 T	Ethyl Acetate	0.767	0.771	-0.5	89	0.00
38 T	Carbon Tetrachloride	0.464	0.495	-6.7	94	0.00
39 T	Methylcyclohexane	0.632	0.691	-9.3	97	0.00
40 TM	Benzene	1.601	1.698	-6.1	92	0.00
41 T	Methacrylonitrile	0.409	0.432	-5.6	87	0.00
42 TM	1,2-Dichloroethane	0.569	0.599	-5.3	93	0.00
43 T	Isopropyl Acetate	1.202	1.251	-4.1	91	0.00
44 TM	Trichloroethene	0.344	0.370	-7.6	95	0.00
45 C	1,2-Dichloropropane	0.462	0.508	-10.0#	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111218\
 Data File : VU028055.D
 Acq On : 12 Nov 2018 11:01
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 13 02:53:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.265	0.285	-7.5	94	0.00
47 T	Bromodichloromethane	0.511	0.574	-12.3	96	0.00
48 T	Methyl methacrylate	0.589	0.626	-6.3	90	0.00
49 T	1,4-Dioxane	0.010	0.009	10.0	83	0.00
50 S	Toluene-d8	1.218	1.329	-9.1	102	0.00
51 T	4-Methyl-2-Pentanone	0.803	0.801	0.2	87	0.00
52 CM	Toluene	0.941	1.014	-7.8#	94	0.00
53 T	t-1,3-Dichloropropene	0.572	0.686	-19.9	98	0.00
54 T	cis-1,3-Dichloropropene	0.631	0.733	-16.2	97	0.00
55 T	1,1,2-Trichloroethane	0.383	0.417	-8.9	97	0.00
56 T	Ethyl methacrylate	0.675	0.729	-8.0	91	0.00
57 T	1,3-Dichloropropane	0.698	0.757	-8.5	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.362	0.392	-8.3	99	0.00
59 T	2-Hexanone	0.621	0.624	-0.5	86	0.00
60 T	Dibromochloromethane	0.362	0.414	-14.4	96	0.00
61 T	1,2-Dibromoethane	0.407	0.444	-9.1	93	0.00
62 S	4-Bromofluorobenzene	0.449	0.502	-11.8	105	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.324	0.353	-9.0	103	0.00
65 PM	Chlorobenzene	1.056	1.112	-5.3	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.386	-6.9	96	0.00
67 C	Ethyl Benzene	1.944	2.079	-6.9#	97	0.00
68 T	m/p-Xylenes	0.706	0.759	-7.5	97	0.00
69 T	o-Xylene	0.712	0.741	-4.1	95	0.00
70 T	Styrene	1.138	1.221	-7.3	94	0.00
71 P	Bromoform	0.302	0.325	-7.6	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.873	3.974	-2.6	97	0.00
74 T	N-amyl acetate	2.342	2.362	-0.9	91	0.00
75 P	1,1,2,2-Tetrachloroethane	1.562	1.520	2.7	92	0.00
76 T	1,2,3-Trichloropropane	1.368	1.282	6.3	91	0.00
77 T	Bromobenzene	0.888	0.904	-1.8	93	0.00
78 T	n-propylbenzene	4.636	4.988	-7.6	99	0.00
79 T	2-Chlorotoluene	2.779	2.829	-1.8	97	0.00
80 T	1,3,5-Trimethylbenzene	3.209	3.366	-4.9	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.507	0.537	-5.9	84	0.00
82 T	4-Chlorotoluene	3.155	3.321	-5.3	100	0.00
83 T	tert-Butylbenzene	3.208	3.218	-0.3	95	0.00
84 T	1,2,4-Trimethylbenzene	3.231	3.442	-6.5	98	0.00
85 T	sec-Butylbenzene	3.901	4.207	-7.8	99	0.00
86 T	p-Isopropyltoluene	3.261	3.579	-9.8	100	0.00
87 T	1,3-Dichlorobenzene	1.641	1.694	-3.2	96	0.00
88 T	1,4-Dichlorobenzene	1.682	1.729	-2.8	98	0.00
89 T	n-Butylbenzene	3.099	3.643	-17.6	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111218\
Data File : VU028055.D
Acq On : 12 Nov 2018 11:01
Operator : MD/SY
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Nov 13 02:53:52 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.581	0.639	-10.0	101	0.00
91 T	1,2-Dichlorobenzene	1.650	1.727	-4.7	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.329	0.319	3.0	87	0.00
93 T	1,2,4-Trichlorobenzene	1.041	1.194	-14.7	98	0.00
94 T	Hexachlorobutadiene	0.524	0.602	-14.9	106	0.00
95 T	Naphthalene	3.725	3.962	-6.4	89	0.00
96 T	1,2,3-Trichlorobenzene	1.107	1.211	-9.4	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6