

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111218\
 Data File : VU028079.D
 Acq On : 12 Nov 2018 20:58
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 13 04:06:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.635	0.511	19.5	65	0.00
3 P	Chloromethane	0.768	0.803	-4.6	90	0.00
4 C	Vinyl Chloride	0.763	0.756	0.9#	85	0.00
5 T	Bromomethane	0.336	0.475	-41.4#	140	0.00
6 T	Chloroethane	0.439	0.459	-4.6	90	0.00
7 T	Trichlorofluoromethane	0.881	0.914	-3.7	89	0.00
8 T	Diethyl Ether	0.377	0.412	-9.3	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.524	0.549	-4.8	89	0.00
10 T	Methyl Iodide	0.392	0.593	-51.3#	122	0.00
11 T	Tert butyl alcohol	0.197	0.220	-11.7	99	0.00
12 CM	1,1-Dichloroethene	0.524	0.511	2.5#	86	0.00
13 T	Acrolein	0.078	0.049	37.2#	64	0.00
14 T	Allyl chloride	1.281	1.287	-0.5	91	0.00
15 T	Acrylonitrile	0.483	0.532	-10.1	93	0.00
16 T	Acetone	0.461	0.472	-2.4	94	0.00
17 T	Carbon Disulfide	1.713	1.594	6.9	81	0.00
18 T	Methyl Acetate	1.269	1.449	-14.2	103	0.00
19 T	Methyl tert-butyl Ether	2.113	2.232	-5.6	92	0.00
20 T	Methylene Chloride	0.673	0.700	-4.0	91	0.00
21 T	trans-1,2-Dichloroethene	0.605	0.590	2.5	85	0.00
22 T	Diisopropyl ether	2.644	2.869	-8.5	94	0.00
23 T	Vinyl Acetate	2.272	2.471	-8.8	91	0.00
24 P	1,1-Dichloroethane	1.317	1.361	-3.3	91	0.00
25 T	2-Butanone	0.734	0.793	-8.0	92	0.00
26 T	2,2-Dichloropropane	0.988	1.017	-2.9	93	0.00
27 T	cis-1,2-Dichloroethene	0.678	0.713	-5.2	88	0.00
28 T	Bromochloromethane	0.617	0.713	-15.6	106	0.00
29 T	Tetrahydrofuran	0.490	0.518	-5.7	92	0.00
30 C	Chloroform	1.208	1.251	-3.6#	92	0.00
31 T	Cyclohexane	1.547	1.307	15.5	90	0.00
32 T	1,1,1-Trichloroethane	0.961	1.026	-6.8	91	0.00
33 S	1,2-Dichloroethane-d4	0.743	0.858	-15.5	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.315	0.358	-13.7	106	0.00
36 T	1,1-Dichloropropene	0.524	0.537	-2.5	91	0.00
37 T	Ethyl Acetate	0.767	0.824	-7.4	94	0.00
38 T	Carbon Tetrachloride	0.464	0.486	-4.7	91	0.00
39 T	Methylcyclohexane	0.632	0.658	-4.1	91	0.00
40 TM	Benzene	1.601	1.689	-5.5	90	0.00
41 T	Methacrylonitrile	0.409	0.462	-13.0	92	0.00
42 TM	1,2-Dichloroethane	0.569	0.600	-5.4	91	0.00
43 T	Isopropyl Acetate	1.202	1.301	-8.2	92	0.00
44 TM	Trichloroethene	0.344	0.363	-5.5	91	0.00
45 C	1,2-Dichloropropane	0.462	0.497	-7.6#	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111218\
 Data File : VU028079.D
 Acq On : 12 Nov 2018 20:58
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 13 04:06:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.265	0.288	-8.7	93	0.00
47 T	Bromodichloromethane	0.511	0.565	-10.6	93	0.00
48 T	Methyl methacrylate	0.589	0.664	-12.7	94	0.00
49 T	1,4-Dioxane	0.010	0.012	-20.0	107	0.00
50 S	Toluene-d8	1.218	1.402	-15.1	106	0.00
51 T	4-Methyl-2-Pentanone	0.803	0.888	-10.6	94	0.00
52 CM	Toluene	0.941	0.990	-5.2#	90	0.00
53 T	t-1,3-Dichloropropene	0.572	0.646	-12.9	91	0.00
54 T	cis-1,3-Dichloropropene	0.631	0.705	-11.7	91	0.00
55 T	1,1,2-Trichloroethane	0.383	0.418	-9.1	96	0.00
56 T	Ethyl methacrylate	0.675	0.753	-11.6	92	0.00
57 T	1,3-Dichloropropane	0.698	0.755	-8.2	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.362	0.406	-12.2	101	0.00
59 T	2-Hexanone	0.621	0.698	-12.4	94	0.00
60 T	Dibromochloromethane	0.362	0.403	-11.3	92	0.00
61 T	1,2-Dibromoethane	0.407	0.443	-8.8	91	0.00
62 S	4-Bromofluorobenzene	0.449	0.532	-18.5	110	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.324	0.337	-4.0	95	0.00
65 PM	Chlorobenzene	1.056	1.101	-4.3	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.389	-7.8	93	0.00
67 C	Ethyl Benzene	1.944	2.068	-6.4#	93	0.00
68 T	m/p-Xylenes	0.706	0.756	-7.1	93	0.00
69 T	o-Xylene	0.712	0.758	-6.5	94	0.00
70 T	Styrene	1.138	1.241	-9.1	93	0.00
71 P	Bromoform	0.302	0.330	-9.3	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	3.873	4.035	-4.2	93	0.00
74 T	N-amyl acetate	2.342	2.616	-11.7	96	0.00
75 P	1,1,2,2-Tetrachloroethane	1.562	1.669	-6.9	96	0.00
76 T	1,2,3-Trichloropropane	1.368	1.400	-2.3	95	0.00
77 T	Bromobenzene	0.888	0.940	-5.9	92	0.00
78 T	n-propylbenzene	4.636	4.942	-6.6	94	0.00
79 T	2-Chlorotoluene	2.779	2.917	-5.0	95	0.00
80 T	1,3,5-Trimethylbenzene	3.209	3.399	-5.9	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.507	0.555	-9.5	82	0.00
82 T	4-Chlorotoluene	3.155	3.326	-5.4	95	0.00
83 T	tert-Butylbenzene	3.208	3.291	-2.6	93	0.00
84 T	1,2,4-Trimethylbenzene	3.231	3.503	-8.4	94	0.00
85 T	sec-Butylbenzene	3.901	4.175	-7.0	93	0.00
86 T	p-Isopropyltoluene	3.261	3.503	-7.4	93	0.00
87 T	1,3-Dichlorobenzene	1.641	1.699	-3.5	92	0.00
88 T	1,4-Dichlorobenzene	1.682	1.703	-1.2	92	0.00
89 T	n-Butylbenzene	3.099	3.287	-6.1	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111218\
Data File : VU028079.D
Acq On : 12 Nov 2018 20:58
Operator : MD/SY
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Nov 13 04:06:53 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.581	0.631	-8.6	95	0.00
91 T	1,2-Dichlorobenzene	1.650	1.762	-6.8	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.329	0.373	-13.4	97	0.00
93 T	1,2,4-Trichlorobenzene	1.041	1.087	-4.4	85	0.00
94 T	Hexachlorobutadiene	0.524	0.530	-1.1	89	0.00
95 T	Naphthalene	3.725	4.050	-8.7	86	0.00
96 T	1,2,3-Trichlorobenzene	1.107	1.134	-2.4	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6