


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU111220\
 Data File : VU041207.D
 Acq On : 12 Nov 2020 09:34
 Operator : SY/MD
 Sample : BFB32
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB202

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUSIM111220.M
 Title : TRACE VOA SOM01.0
 Last Update : Fri Nov 13 09:30:10 2020



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2875

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.4	23157	PASS
75	95	30	80	55.7	60325	PASS
95	95	100	100	100.0	108384	PASS
96	95	5	9	6.7	7266	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	77.1	83589	PASS
175	174	5	9	7.5	6235	PASS
176	174	95	101	95.3	79648	PASS
177	176	5	9	6.3	5054	PASS

m/z	Abundance
36.00	1237.0
37.10	6002.0
38.00	5638.0
39.10	2796.0
39.90	195.0
40.90	100.0
44.00	1003.0
45.10	1049.0
46.00	155.0
47.10	1428.0
48.00	761.0
49.10	5486.0
50.10	23232.0
51.10	7459.0
52.00	455.0
53.10	201.0
55.00	362.0
56.00	1713.0
57.10	3299.0
57.90	128.0
60.00	1267.0
61.00	6185.0
62.10	6582.0
63.10	5865.0
64.00	1094.0
65.00	1203.0
66.10	161.0
66.90	250.0
68.10	11812.0
69.10	12295.0
70.00	902.0
72.10	696.0
73.00	5664.0
74.10	20552.0
75.10	60424.0
76.10	5529.0
77.00	557.0
78.00	320.0
79.00	5191.0
80.00	1536.0
81.00	5188.0
81.90	1180.0
82.90	132.0
85.90	176.0
87.00	4717.0
88.10	3937.0
91.10	481.0
92.10	3358.0
93.10	5333.0
94.10	14146.0
95.10	107296.0
96.10	7404.0
97.00	244.0
104.00	557.0
105.00	208.0
105.90	642.0
106.90	128.0
108.00	125.0
111.00	195.0
112.80	134.0
116.00	537.0
117.00	1097.0
118.00	503.0
119.00	1032.0
128.00	406.0
128.90	157.0
130.00	416.0
131.00	197.0
135.00	206.0
137.00	229.0
141.00	1422.0
141.90	150.0
143.00	1939.0
144.00	148.0
145.00	769.0
146.00	189.0
148.00	282.0
150.00	125.0
155.00	240.0
157.00	212.0
159.00	119.0
171.00	367.0
172.10	1183.0
174.10	80648.0
175.10	6041.0
176.10	77088.0
177.10	5007.0
178.10	107.0

Instrument :
MSVOA_U
ClientSampleId :
BFB202

m/z	Abundance
36.00	1211.0
37.10	6155.0
38.10	5751.0
39.10	2745.0
40.00	187.0
44.00	1029.0
45.10	1111.0
46.00	173.0
47.10	1472.0
48.00	830.0
49.10	5661.0
50.10	24024.0
51.10	7766.0
52.00	422.0
53.00	220.0
54.90	333.0
56.00	1772.0
57.10	3476.0
57.90	110.0
60.00	1274.0
61.00	6346.0
62.10	6489.0
63.10	5831.0
64.00	1042.0
65.10	1216.0
66.00	132.0
67.00	284.0
68.10	12332.0
69.10	12850.0
70.00	937.0
72.10	697.0
73.00	5988.0
74.10	21352.0
75.10	62440.0
76.10	5615.0
77.10	567.0
78.00	265.0
79.00	5281.0
80.00	1539.0
81.00	5316.0
82.00	1193.0
82.90	128.0
85.90	166.0
87.10	4737.0
88.10	4245.0
91.00	486.0
92.10	3420.0
93.10	5724.0
94.10	14442.0
95.10	112200.0
96.10	7543.0
97.00	236.0
104.00	603.0
105.00	189.0
105.90	646.0
107.00	166.0
108.00	144.0
111.00	201.0
112.80	135.0
114.90	111.0
116.00	552.0
117.00	1052.0
118.00	537.0
119.00	927.0
128.00	429.0
129.00	172.0
130.00	449.0
131.00	244.0
135.00	214.0
136.90	277.0
141.00	1580.0
142.00	149.0
143.10	2049.0
144.00	145.0
145.00	774.0
145.90	195.0
148.10	295.0
149.00	112.0
150.00	106.0
155.00	268.0
157.00	219.0
160.90	128.0
171.00	372.0
172.10	1204.0
174.10	86456.0
175.10	6466.0
176.10	82328.0
177.10	5213.0
178.00	103.0

Instrument :
MSVOA_U
ClientSampleId :
BFB202

m/z	Abundance
36.00	1084.0
37.00	5653.0
38.00	5160.0
39.00	2400.0
40.00	188.0
44.00	914.0
45.00	1051.0
46.00	167.0
47.00	1233.0
48.00	776.0
49.10	5034.0
50.10	22216.0
51.10	7043.0
52.00	339.0
53.00	203.0
54.90	295.0
56.00	1724.0
57.10	3334.0
60.00	1137.0
61.10	5967.0
62.10	5824.0
63.10	5347.0
64.10	857.0
65.10	1050.0
65.90	123.0
68.10	11451.0
69.10	12166.0
70.00	881.0
72.10	642.0
73.10	5686.0
74.10	20112.0
75.10	58112.0
76.10	5107.0
77.10	591.0
77.80	248.0
79.00	4999.0
80.00	1341.0
81.00	5071.0
82.00	1114.0
85.90	153.0
87.00	4159.0
88.10	4098.0
91.00	478.0
92.10	3211.0
93.10	5369.0
94.10	13651.0
95.10	105656.0
96.10	6852.0
97.00	199.0
104.00	570.0
104.90	158.0
105.90	608.0
107.00	172.0
108.00	102.0
111.00	155.0
112.90	142.0
114.80	131.0
115.90	513.0
117.00	888.0
118.00	561.0
119.00	776.0
128.00	408.0
129.10	172.0
130.00	450.0
131.00	229.0
135.00	205.0
136.90	355.0
141.00	1504.0
142.00	147.0
143.10	1980.0
144.00	133.0
145.00	712.0
146.00	174.0
147.00	102.0
148.10	268.0
155.00	275.0
157.00	207.0
160.90	131.0
171.00	304.0
172.10	1064.0
174.10	83664.0
175.10	6198.0
176.10	79528.0
177.10	4944.0
178.00	132.0

Instrument :
MSVOA_U
ClientSampleId :
BFB202

m/z	Abundance
44.00	201.0

Instrument :
MSVOA_U
ClientSampleId :
BFB202