

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
 Data File : VU045757.D
 Acq On : 12 Nov 2021 10:22
 Operator : SY/MD
 Sample : VU1112MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1112MBL01

Quant Time: Nov 15 05:16:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	115164	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	211534	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	211112	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	95910	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	93424	54.168	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.340%
35) Dibromofluoromethane	5.295	113	75031	53.127	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.260%
50) Toluene-d8	7.902	98	277112	52.767	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.540%
62) 4-Bromofluorobenzene	10.635	95	97107	46.112	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
Data File : VU045757.D
Acq On : 12 Nov 2021 10:22
Operator : SY/MD
Sample : VU1112MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1112MBL01

Quant Time: Nov 15 05:16:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

