

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
 Data File : VU045758.D
 Acq On : 12 Nov 2021 10:45
 Operator : SY/MD
 Sample : VU1112WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1112WBL01

Quant Time: Nov 15 05:17:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	117203	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	219124	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	225528	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	105626	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	96650	55.063	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.120%
35) Dibromofluoromethane	5.292	113	76920	52.578	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.160%
50) Toluene-d8	7.899	98	287683	52.882	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.760%
62) 4-Bromofluorobenzene	10.635	95	106265	48.713	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
Data File : VU045758.D
Acq On : 12 Nov 2021 10:45
Operator : SY/MD
Sample : VU1112WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1112WBL01

Quant Time: Nov 15 05:17:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

