

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
 Data File : VU045761.D
 Acq On : 12 Nov 2021 12:06
 Operator : SY/MD
 Sample : M4606-19 40X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP-01-20211109

Quant Time: Nov 15 05:18:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	116138	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	213968	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	215678	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	102888	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	92117	52.962	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.920%
35) Dibromofluoromethane	5.292	113	73725	51.608	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.220%
50) Toluene-d8	7.902	98	278559	52.439	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.880%
62) 4-Bromofluorobenzene	10.635	95	103599	48.636	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
Data File : VU045761.D
Acq On : 12 Nov 2021 12:06
Operator : SY/MD
Sample : M4606-19 40X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP-01-20211109

Quant Time: Nov 15 05:18:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

