

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
 Data File : VU045766.D
 Acq On : 12 Nov 2021 14:02
 Operator : SY/MD
 Sample : M4606-20
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP-02-20211109

Quant Time: Nov 15 05:19:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	117694	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	219125	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	228910	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	112356	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	95912	54.415	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.840%	
35) Dibromofluoromethane	5.292	113	75098	51.332	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.660%	
50) Toluene-d8	7.899	98	289883	53.287	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.580%	
62) 4-Bromofluorobenzene	10.635	95	108278	49.636	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.280%	
Target Compounds						
16) Acetone	2.629	43	1664	1.855	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
Data File : VU045766.D
Acq On : 12 Nov 2021 14:02
Operator : SY/MD
Sample : M4606-20
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP-02-20211109

Quant Time: Nov 15 05:19:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

