

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
 Data File : VU045767.D
 Acq On : 12 Nov 2021 14:25
 Operator : SY/MD
 Sample : M4668-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW9-DUP-20211109

Quant Time: Nov 15 05:19:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.376	168	114476	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	212394	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	219967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	109750	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	95555	55.736	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.480%
35) Dibromofluoromethane	5.292	113	73496	51.829	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.660%
50) Toluene-d8	7.900	98	281607	53.406	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.820%
62) 4-Bromofluorobenzene	10.636	95	106167	50.210	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.420%
Target Compounds						
16) Acetone	2.630	43	2590	2.968	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
Data File : VU045767.D
Acq On : 12 Nov 2021 14:25
Operator : SY/MD
Sample : M4668-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW9-DUP-20211109

Quant Time: Nov 15 05:19:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

