

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
 Data File : VU045768.D
 Acq On : 12 Nov 2021 14:48
 Operator : SY/MD
 Sample : M4668-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW9-GW-658-660

Quant Time: Nov 15 05:20:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	117376	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	214039	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	225803	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	112421	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	94832	53.948	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.900%	
35) Dibromofluoromethane	5.292	113	74294	51.989	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.980%	
50) Toluene-d8	7.903	98	286649	53.944	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.880%	
62) 4-Bromofluorobenzene	10.636	95	107734	50.560	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.120%	
Target Compounds						
16) Acetone	2.626	43	2278	2.546	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
Data File : VU045768.D
Acq On : 12 Nov 2021 14:48
Operator : SY/MD
Sample : M4668-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW9-GW-658-660

Quant Time: Nov 15 05:20:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

