

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
 Data File : VU045774.D
 Acq On : 12 Nov 2021 17:07
 Operator : SY/MD
 Sample : M4687-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TRIP-BLANK

Quant Time: Nov 15 05:21:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	117022	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	216606	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	227004	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	113792	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	97753	55.778	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.560%
35) Dibromofluoromethane	5.292	113	75554	52.244	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.480%
50) Toluene-d8	7.899	98	290263	53.977	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.960%
62) 4-Bromofluorobenzene	10.636	95	110300	51.151	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.300%
Target Compounds						
						Qvalue
3) Chloromethane	1.523	50	634	0.365	ug/l	91
16) Acetone	2.626	43	1622	1.818	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
Data File : VU045774.D
Acq On : 12 Nov 2021 17:07
Operator : SY/MD
Sample : M4687-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 15 05:21:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

