

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
 Data File : VU045780.D
 Acq On : 12 Nov 2021 19:26
 Operator : SY/MD
 Sample : M4644-09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 G-11

Quant Time: Nov 15 05:24:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	117123	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	221535	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	231381	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	108833	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	96868	55.225	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.460%
35) Dibromofluoromethane	5.292	113	75833	51.271	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.540%
50) Toluene-d8	7.899	98	296249	53.864	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.720%
62) 4-Bromofluorobenzene	10.636	95	108811	49.338	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.680%
Target Compounds						Qvalue
16) Acetone	2.626	43	24040	26.924	ug/l	98
20) Methylene Chloride	3.047	84	8912	5.744	ug/l	94
43) Isopropyl Acetate	5.916	43	14819	3.905	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
Data File : VU045780.D
Acq On : 12 Nov 2021 19:26
Operator : SY/MD
Sample : M4644-09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
G-11

Quant Time: Nov 15 05:24:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

