

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
 Data File : VU045781.D
 Acq On : 12 Nov 2021 19:49
 Operator : SY/MD
 Sample : M4640-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-2-AS

Quant Time: Nov 15 05:24:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	114458	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	220312	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	230244	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	113157	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	95093	55.476	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 110.960%	
35) Dibromofluoromethane	5.292	113	74716	50.796	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.600%	
50) Toluene-d8	7.899	98	295248	53.980	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.960%	
62) 4-Bromofluorobenzene	10.636	95	110970	50.596	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.200%	
Target Compounds						Qvalue
16) Acetone	2.626	43	24407	27.971	ug/l	97
20) Methylene Chloride	3.051	84	9739	6.423	ug/l	96
43) Isopropyl Acetate	5.916	43	17011	4.507	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111221\
Data File : VU045781.D
Acq On : 12 Nov 2021 19:49
Operator : SY/MD
Sample : M4640-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-2-AS

Quant Time: Nov 15 05:24:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

