

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111225\
 Data File : VU063710.D
 Acq On : 12 Nov 2025 10:22
 Operator : MD/SY
 Sample : VU1112WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1112WBL01

Quant Time: Nov 14 04:41:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U111025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 01:41:19 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.345	168	181067	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.222	114	282529	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.393	117	263432	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.788	152	141364	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.673	65	150700	56.341	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.680%
35) Dibromofluoromethane	5.258	113	119795	60.538	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	121.080%
50) Toluene-d8	7.872	98	376370	52.943	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.880%
62) 4-Bromofluorobenzene	10.608	95	171146	58.198	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111225\
Data File : VU063710.D
Acq On : 12 Nov 2025 10:22
Operator : MD/SY
Sample : VU1112WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1112WBL01

Quant Time: Nov 14 04:41:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U111025W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 11 01:41:19 2025
Response via : Initial Calibration

