

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111318\
 Data File : VU028089.D
 Acq On : 13 Nov 2018 14:05
 Operator : MD/SY
 Sample : J5885-01DL 20X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RW-2DL

Quant Time: Nov 14 09:39:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	264874	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	486298	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	430775	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	182695	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	200536	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
35) Dibromofluoromethane	4.89	113	154612	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
50) Toluene-d8	7.57	98	616602	52.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.12%	
62) 4-Bromofluorobenzene	10.31	95	205706	47.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.18%	
Target Compounds						
64) Tetrachloroethene	8.23	164	102862	36.825	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU111318\
Data File : VU028089.D
Acq On : 13 Nov 2018 14:05
Operator : MD/SY
Sample : J5885-01DL 20X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RW-2DL

Quant Time: Nov 14 09:39:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

