

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111318\
 Data File : VU028090.D
 Acq On : 13 Nov 2018 14:29
 Operator : MD/SY
 Sample : J5885-03DL 100X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RW-1DL

Quant Time: Nov 14 09:39:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	269491	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	484780	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	423184	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	169764	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	200923	50.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.30%	
35) Dibromofluoromethane	4.88	113	155565	50.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.78%	
50) Toluene-d8	7.57	98	626672	53.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.14%	
62) 4-Bromofluorobenzene	10.31	95	198471	45.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.14%	
Target Compounds						
21) trans-1,2-Dichloroethene	2.98	96	3802	1.166	ug/l	95
27) cis-1,2-Dichloroethene	4.23	96	499702	136.712	ug/l	86
44) Trichloroethene	6.19	130	148623	44.541	ug/l	92
64) Tetrachloroethene	8.23	164	176633	64.370	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU111318\
Data File : VU028090.D
Acq On : 13 Nov 2018 14:29
Operator : MD/SY
Sample : J5885-03DL 100X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RW-1DL

Quant Time: Nov 14 09:39:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

