

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111318\
 Data File : VU028091.D
 Acq On : 13 Nov 2018 14:53
 Operator : MD/SY
 Sample : J5885-04DL2 500X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-2DL2

Quant Time: Nov 14 09:40:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	272743	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	496455	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	436279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	174060	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	209500	51.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.32%	
35) Dibromofluoromethane	4.88	113	160082	51.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.28%	
50) Toluene-d8	7.57	98	637612	52.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.46%	
62) 4-Bromofluorobenzene	10.31	95	204782	45.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.84%	
Target Compounds						
4) Vinyl Chloride	1.41	62	88856	21.359	ug/l	99
27) cis-1,2-Dichloroethene	4.23	96	106492	28.787	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU111318\
Data File : VU028091.D
Acq On : 13 Nov 2018 14:53
Operator : MD/SY
Sample : J5885-04DL2 500X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-2DL2

Quant Time: Nov 14 09:40:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

