

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111318\
 Data File : VU028092.D
 Acq On : 13 Nov 2018 15:18
 Operator : MD/SY
 Sample : J5885-06DL2 1000X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-5DL2

Quant Time: Nov 14 09:41:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	262542	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	480809	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	414184	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	163838	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	202371	51.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.70%	
35) Dibromofluoromethane	4.89	113	154192	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
50) Toluene-d8	7.57	98	612733	52.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.64%	
62) 4-Bromofluorobenzene	10.31	95	193972	44.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.82%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.23	96	8995	2.526	ug/l	80
64) Tetrachloroethene	8.23	164	89633	33.374	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU111318\
Data File : VU028092.D
Acq On : 13 Nov 2018 15:18
Operator : MD/SY
Sample : J5885-06DL2 1000X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-5DL2

Quant Time: Nov 14 09:41:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

