

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111318\
 Data File : VU028093.D
 Acq On : 13 Nov 2018 15:42
 Operator : MD/SY
 Sample : J5885-07DL2 2000X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-4DL2

Quant Time: Nov 14 09:41:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	260867	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	470943	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	424458	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	172563	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	199938	51.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.10%	
35) Dibromofluoromethane	4.88	113	151716	51.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.18%	
50) Toluene-d8	7.57	98	606605	52.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.76%	
62) 4-Bromofluorobenzene	10.31	95	199342	47.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.24%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.24	96	3562	1.007	ug/l	89
64) Tetrachloroethene	8.23	164	99226	36.052	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU111318\
Data File : VU028093.D
Acq On : 13 Nov 2018 15:42
Operator : MD/SY
Sample : J5885-07DL2 2000X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-4DL2

Quant Time: Nov 14 09:41:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

