

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111318\
 Data File : VU028099.D
 Acq On : 13 Nov 2018 18:06
 Operator : MD/SY
 Sample : J5909-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SB-15-S

Quant Time: Nov 14 08:00:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	267992	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	488392	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	430360	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	171014	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	202331	50.78	ug/l	0.00
Spiked Amount			Recovery	=	101.56%	
35) Dibromofluoromethane	4.88	113	156148	50.71	ug/l	0.00
Spiked Amount			Recovery	=	101.42%	
50) Toluene-d8	7.57	98	618438	51.99	ug/l	0.00
Spiked Amount			Recovery	=	103.98%	
62) 4-Bromofluorobenzene	10.31	95	199991	45.58	ug/l	0.00
Spiked Amount			Recovery	=	91.16%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.32	43	36672	14.835	ug/l	98
18) Methyl Acetate	2.62	43	8518	1.253	ug/l	100
20) Methylene Chloride	2.70	84	20799	5.763	ug/l	95
25) 2-Butanone	4.29	43	8469	2.153	ug/l	92
43) Isopropyl Acetate	5.55	43	252970	21.544	ug/l	97
95) Naphthalene	13.75	128	22382	1.419	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU111318\
Data File : VU028099.D
Acq On : 13 Nov 2018 18:06
Operator : MD/SY
Sample : J5909-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SB-15-S

Quant Time: Nov 14 08:00:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

