

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111318\
 Data File : VU028105.D
 Acq On : 13 Nov 2018 20:31
 Operator : MD/SY
 Sample : J5908-08
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SR-3-S

Quant Time: Nov 14 08:09:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	248307	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	454065	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	398924	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	164177	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	190290	51.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.10%	
35) Dibromofluoromethane	4.88	113	144326	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
50) Toluene-d8	7.57	98	580317	52.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.94%	
62) 4-Bromofluorobenzene	10.31	95	192048	47.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.16%	
Target Compounds						
16) Acetone	2.32	43	18224	7.956	ug/l #	89
20) Methylene Chloride	2.71	84	14802	4.426	ug/l	94
43) Isopropyl Acetate	5.56	43	25008	2.291	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU111318\
Data File : VU028105.D
Acq On : 13 Nov 2018 20:31
Operator : MD/SY
Sample : J5908-08
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SR-3-S

Quant Time: Nov 14 08:09:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

