

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111318\
 Data File : VU028106.D
 Acq On : 13 Nov 2018 20:55
 Operator : MD/SY
 Sample : J5908-12
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SR-4-S

Quant Time: Nov 14 08:10:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	247231	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	447394	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	404191	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	165922	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	192908	52.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.96%	
35) Dibromofluoromethane	4.88	113	145796	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	
50) Toluene-d8	7.57	98	575963	52.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.72%	
62) 4-Bromofluorobenzene	10.31	95	192968	48.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.02%	
Target Compounds						
16) Acetone	2.32	43	17945	7.869	ug/l	100
20) Methylene Chloride	2.70	84	14167	4.255	ug/l	91
43) Isopropyl Acetate	5.56	43	25040	2.328	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU111318\
Data File : VU028106.D
Acq On : 13 Nov 2018 20:55
Operator : MD/SY
Sample : J5908-12
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SR-4-S

Quant Time: Nov 14 08:10:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

