

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111318\
 Data File : VU028107.D
 Acq On : 13 Nov 2018 21:19
 Operator : MD/SY
 Sample : J5908-16
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SR-5-S

Quant Time: Nov 14 08:12:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	238040	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	439502	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	384305	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	157511	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	183227	51.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.54%	
35) Dibromofluoromethane	4.88	113	138868	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.22%	
50) Toluene-d8	7.57	98	557293	52.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.12%	
62) 4-Bromofluorobenzene	10.31	95	184712	46.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.56%	
Target Compounds						
16) Acetone	2.32	43	16479	7.505	ug/l	93
18) Methyl Acetate	2.62	43	11422	1.891	ug/l	100
20) Methylene Chloride	2.70	84	23292	7.265	ug/l	98
43) Isopropyl Acetate	5.55	43	242331	22.934	ug/l	99
95) Naphthalene	13.75	128	19559	1.337	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU111318\
Data File : VU028107.D
Acq On : 13 Nov 2018 21:19
Operator : MD/SY
Sample : J5908-16
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SR-5-S

Quant Time: Nov 14 08:12:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

