

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
 Data File : VU056303.D
 Acq On : 14 Nov 2023 14:13
 Operator : MD/SY
 Sample : 05341-05
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0QJ5

Quant Time: Nov 15 02:54:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 15 01:36:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	323528	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	319915	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	144048	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	61617	2.546	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	51.000%
7) Chloroethane-d5	1.907	69	53400	2.878	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	57.600%#
11) 1,1-Dichloroethene-d2	2.563	65	31160	2.918	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.400%#
20) 2-Butanone-d5	4.621	46	175008	36.269	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	72.540%
24) Chloroform-d	5.055	84	155686	3.326	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	66.600%#
26) 1,2-Dichloroethane-d4	5.698	65	74060	3.424	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	68.400%#
32) Benzene-d6	5.721	84	298858	3.135	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	62.800%#
36) 1,2-Dichloropropane-d6	6.685	67	95693	3.331	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	66.600%
41) Toluene-d8	7.894	98	256043	3.042	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	60.800%#
43) trans-1,3-Dichloroprop...	8.177	79	33544	3.045	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	60.800%
46) 2-Hexanone-d5	8.631	63	144651	33.410	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	66.820%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	70568	3.438	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	68.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	93593	3.517	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	70.400%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
Data File : VU056303.D
Acq On : 14 Nov 2023 14:13
Operator : MD/SY
Sample : 05341-05
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0QJ5

Quant Time: Nov 15 02:54:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 15 01:36:14 2023
Response via : Initial Calibration

