

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
 Data File : VU056307.D
 Acq On : 14 Nov 2023 15:50
 Operator : MD/SY
 Sample : 05341-09
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0QK4

Quant Time: Nov 15 02:54:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 15 01:36:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	291920	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	287384	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	128394	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	59766	2.737	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	54.800%
7) Chloroethane-d5	1.907	69	53688	3.207	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	64.200%#
11) 1,1-Dichloroethene-d2	2.563	65	31674	3.287	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.800%
20) 2-Butanone-d5	4.618	46	172739	39.675	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.340%
24) Chloroform-d	5.055	84	152327	3.607	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	72.200%
26) 1,2-Dichloroethane-d4	5.698	65	75598	3.874	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.400%
32) Benzene-d6	5.721	84	293162	3.424	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	68.400%#
36) 1,2-Dichloropropane-d6	6.685	67	93475	3.622	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	72.400%
41) Toluene-d8	7.894	98	247306	3.271	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	65.400%#
43) trans-1,3-Dichloroprop...	8.177	79	32084	3.242	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	64.800%
46) 2-Hexanone-d5	8.631	63	148495	38.180	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.360%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	70661	3.832	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	76.600%
66) 1,2-Dichlorobenzene-d4	12.190	152	90436	3.813	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	76.200%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
Data File : VU056307.D
Acq On : 14 Nov 2023 15:50
Operator : MD/SY
Sample : 05341-09
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0QK4

Quant Time: Nov 15 02:54:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 15 01:36:14 2023
Response via : Initial Calibration

