

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
 Data File : VU056311.D
 Acq On : 14 Nov 2023 17:28
 Operator : MD/SY
 Sample : 05328-12
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A49W0

Quant Time: Nov 15 02:55:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 15 01:36:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	280718	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	274857	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	122709	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	59870	2.851	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	57.000%
7) Chloroethane-d5	1.907	69	58403	3.628	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	72.600%
11) 1,1-Dichloroethene-d2	2.563	65	33625	3.629	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.600%
20) 2-Butanone-d5	4.618	46	159766	38.160	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	76.320%
24) Chloroform-d	5.058	84	157575	3.880	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.600%
26) 1,2-Dichloroethane-d4	5.698	65	75573	4.027	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.600%
32) Benzene-d6	5.721	84	297304	3.630	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	72.600%
36) 1,2-Dichloropropane-d6	6.685	67	97771	3.961	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.200%
41) Toluene-d8	7.894	98	251945	3.484	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.600%#
43) trans-1,3-Dichloroprop...	8.177	79	33158	3.503	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	70.000%
46) 2-Hexanone-d5	8.627	63	134420	36.136	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.280%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	69651	3.949	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	93384	4.119	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	82.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
Data File : VU056311.D
Acq On : 14 Nov 2023 17:28
Operator : MD/SY
Sample : 05328-12
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A49W0

Quant Time: Nov 15 02:55:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 15 01:36:14 2023
Response via : Initial Calibration

