

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
 Data File : VU056315.D
 Acq On : 14 Nov 2023 19:05
 Operator : MD/SY
 Sample : 05333-06
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YCG81

Quant Time: Nov 15 02:56:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 15 01:36:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	282033	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	283357	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	122401	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	56447	2.676	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	53.600%
7) Chloroethane-d5	1.908	69	53486	3.307	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	66.200%
11) 1,1-Dichloroethene-d2	2.563	65	28946	3.109	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.200%
20) 2-Butanone-d5	4.621	46	165086	39.246	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.500%
24) Chloroform-d	5.059	84	146763	3.597	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	72.000%
26) 1,2-Dichloroethane-d4	5.695	65	73000	3.872	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.400%
32) Benzene-d6	5.721	84	279489	3.311	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	66.200%#
36) 1,2-Dichloropropane-d6	6.685	67	92047	3.618	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	72.400%
41) Toluene-d8	7.894	98	240420	3.225	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	64.400%#
43) trans-1,3-Dichloroprop...	8.177	79	32652	3.346	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	67.000%
46) 2-Hexanone-d5	8.631	63	138028	35.993	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	71.980%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	68388	3.761	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	75.200%
66) 1,2-Dichlorobenzene-d4	12.190	152	88853	3.929	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	78.600%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
Data File : VU056315.D
Acq On : 14 Nov 2023 19:05
Operator : MD/SY
Sample : 05333-06
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YCG81

Quant Time: Nov 15 02:56:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 15 01:36:14 2023
Response via : Initial Calibration

