

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
 Data File : VU056316.D
 Acq On : 14 Nov 2023 19:29
 Operator : MD/SY
 Sample : 05333-07
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YCG83

Quant Time: Nov 15 02:56:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 15 01:36:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	280303	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	279860	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	123851	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	58362	2.784	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	55.600%
7) Chloroethane-d5	1.907	69	55111	3.428	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	68.600%
11) 1,1-Dichloroethene-d2	2.563	65	31295	3.382	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.600%
20) 2-Butanone-d5	4.618	46	169491	40.542	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.080%
24) Chloroform-d	5.055	84	149752	3.693	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.800%
26) 1,2-Dichloroethane-d4	5.695	65	73884	3.943	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.800%
32) Benzene-d6	5.721	84	281940	3.381	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	67.600%#
36) 1,2-Dichloropropane-d6	6.685	67	91210	3.630	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	72.600%
41) Toluene-d8	7.894	98	240627	3.268	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	65.400%#
43) trans-1,3-Dichloroprop...	8.177	79	31959	3.316	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	66.400%
46) 2-Hexanone-d5	8.630	63	139372	36.798	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	73.600%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	69481	3.869	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	77.400%
66) 1,2-Dichlorobenzene-d4	12.190	152	88574	3.871	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	77.400%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
Data File : VU056316.D
Acq On : 14 Nov 2023 19:29
Operator : MD/SY
Sample : 05333-07
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YCG83

Quant Time: Nov 15 02:56:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 15 01:36:14 2023
Response via : Initial Calibration

