

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
 Data File : VU056260.D
 Acq On : 13 Nov 2023 18:53
 Operator : MD/SY
 Sample : VU1113WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK069

Quant Time: Nov 13 23:54:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 13 23:42:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	323932	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	313212	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	138132	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	114614	4.730	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.600%
7) Chloroethane-d5	1.908	69	91697	4.936	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.800%
11) 1,1-Dichloroethene-d2	2.563	65	50794	4.750	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	95.000%
20) 2-Butanone-d5	4.618	46	230038	47.614	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.220%
24) Chloroform-d	5.055	84	215897	4.607	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.200%
26) 1,2-Dichloroethane-d4	5.695	65	106982	4.940	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.800%
32) Benzene-d6	5.721	84	442188	4.738	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.800%
36) 1,2-Dichloropropane-d6	6.685	67	135761	4.827	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.600%
41) Toluene-d8	7.894	98	374792	4.548	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.000%
43) trans-1,3-Dichloroprop...	8.177	79	50937	4.722	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.400%
46) 2-Hexanone-d5	8.628	63	198898	46.922	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.840%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	93211	4.638	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	126815	4.970	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
Data File : VU056260.D
Acq On : 13 Nov 2023 18:53
Operator : MD/SY
Sample : VU1113WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK069

Quant Time: Nov 13 23:54:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Nov 13 23:42:10 2023
Response via : Initial Calibration

