

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
 Data File : VU056274.D
 Acq On : 14 Nov 2023 00:32
 Operator : MD/SY
 Sample : 05372-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCDP9

Quant Time: Nov 14 21:47:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 14 00:08:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	305713	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	302760	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	135428	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	67752	2.963	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	59.200%
7) Chloroethane-d5	1.911	69	59796	3.411	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	68.200%
11) 1,1-Dichloroethene-d2	2.563	65	35456	3.513	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.200%
20) 2-Butanone-d5	4.631	46	188808	41.409	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.820%
24) Chloroform-d	5.055	84	160109	3.620	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	72.400%
26) 1,2-Dichloroethane-d4	5.698	65	77141	3.774	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.400%
32) Benzene-d6	5.721	84	313089	3.471	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	69.400%#
36) 1,2-Dichloropropane-d6	6.685	67	99760	3.669	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	73.400%
41) Toluene-d8	7.894	98	266660	3.348	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	67.000%#
43) trans-1,3-Dichloroprop...	8.177	79	35777	3.431	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	68.600%
46) 2-Hexanone-d5	8.631	63	155029	37.835	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.680%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	72427	3.728	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	74.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	97029	3.878	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	77.600%#
Target Compounds						
					Qvalue	
13) Acetone	2.640	43	31726	8.170	ug/L	97
34) Trichloroethene	6.541	95	75039	3.188	ug/L	95
47) Tetrachloroethene	8.547	164	4607	0.290	ug/L	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
Data File : VU056274.D
Acq On : 14 Nov 2023 00:32
Operator : MD/SY
Sample : 05372-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCDP9

Quant Time: Nov 14 21:47:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 14 00:08:30 2023
Response via : Initial Calibration





#13

Acetone

Concen: 8.170 ug/L

RT: 2.640 min Scan# 41

Delta R.T. 0.010 min

Lab File: VU056274.D

Acq: 14 Nov 2023 00:32

Instrument :

MSVOA_U

ClientSampleId :





GCDP9

Tgt Ion: 43 Resp: 31726

Ion Ratio Lower Upper

43 100

58 36.8 0.0 70.0



#34

Trichloroethene

Concen: 3.188 ug/L

RT: 6.541 min Scan# 1633

Delta R.T. 0.003 min

Lab File: VU056274.D

Acq: 14 Nov 2023 00:32

Tgt Ion: 95 Resp: 75039




Ion Ratio Lower Upper


95 100

97 61.7 46.5 86.3

132 90.3 62.0 115.2

130 88.7 66.3 123.1





#47
 Tetrachloroethene
 Concen: 0.290 ug/L
 RT: 8.547 min Scan# 21
 Delta R.T. -0.003 min
 Lab File: VU056274.D
 Acq: 14 Nov 2023 00:32

Instrument :
 MSVOA_U
 ClientSampleId :
 GCDP9

Tgt Ion:164 Resp: 4607

Ion	Ratio	Lower	Upper
164	100		
129	97.6	71.0	131.9
131	96.3	70.3	130.7
166	110.4	89.8	166.8

