

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
 Data File : VU056305.D
 Acq On : 14 Nov 2023 15:02
 Operator : MD/SY
 Sample : 05341-07
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0QJ8

Quant Time: Nov 15 02:54:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 15 01:36:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	292149	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	286541	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	128133	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	60941	2.789	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	55.800%
7) Chloroethane-d5	1.907	69	53873	3.215	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	64.400%#
11) 1,1-Dichloroethene-d2	2.563	65	33585	3.483	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.600%
20) 2-Butanone-d5	4.621	46	172298	39.543	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.080%
24) Chloroform-d	5.055	84	151728	3.590	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	71.800%
26) 1,2-Dichloroethane-d4	5.695	65	74913	3.836	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.800%
32) Benzene-d6	5.724	84	289495	3.391	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	67.800%#
36) 1,2-Dichloropropane-d6	6.685	67	92591	3.599	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	72.000%
41) Toluene-d8	7.894	98	247878	3.288	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	65.800%#
43) trans-1,3-Dichloroprop...	8.177	79	32922	3.336	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	66.800%
46) 2-Hexanone-d5	8.631	63	141918	36.596	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	73.200%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	67977	3.697	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	74.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	92557	3.910	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	78.200%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
Data File : VU056305.D
Acq On : 14 Nov 2023 15:02
Operator : MD/SY
Sample : 05341-07
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0QJ8

Quant Time: Nov 15 02:54:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 15 01:36:14 2023
Response via : Initial Calibration

